



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB130122

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/02/2024	07:45
2	CAL2	1	Water	NA	NA	20.3	7.00	04/02/2024	07:46
3	CAL3	1	Water	NA	NA	20.3	10.02	04/02/2024	07:49
4	ICV	1	Water	NA	NA	20.3	7.01	04/02/2024	07:50
5	CCV1	1	Water	NA	NA	20.3	2.01	04/02/2024	08:00
6	P1915-02	1	Solid	20.02	20	24.2	9.83	04/02/2024	08:11
7	P1927-02	1	Solid	20.03	20	23.9	7.24	04/02/2024	08:20
8	P1927-04	1	Solid	20.01	20	23.3	8.41	04/02/2024	08:25
9	P1927-04DUP	1	Solid	20.02	20	23.4	8.42	04/02/2024	08:26
10	CCV2	1	Water	NA	NA	20.3	12.02	04/02/2024	08:33

130122

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY P1927

WorkList ID : 179030

Department : Wet-Chemistry

Date : 04-02-2024 07:34:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1915-02	WC-A-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	I22	03/28/2024	9045D
P1927-02	72-11991	Solid	Corrosivity	Cool 4 deg C	PSEG03	I32	03/29/2024	9045D
P1927-04	72-11998	Solid	Corrosivity	Cool 4 deg C	PSEG03	I32	03/29/2024	9045D

Date/Time 04/02/24 07:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04/02/24 10:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]