



Analytical Summary Report

Analysis Method: 120.1

ANALYST: Iwona

Parameter: Conductance

SUPERVISOR REVIEW BY: Sohil

Run Number: LB130126

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP104779

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB130126BL		1	24.2	0.82	0.83	04/02/2024	11:10
2	LB130126BS	1413	1	24.7	1399.00	1407.06	04/02/2024	11:15
3	P1911-08		1	24.6	323.00	325.49	04/02/2024	11:20

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$