

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB130144

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/03/2024	11:45
2	CAL2	1	Water	NA	NA	20.3	7.00	04/03/2024	11:46
3	CAL3	1	Water	NA	NA	20.3	10.03	04/03/2024	11:47
4	ICV	1	Water	NA	NA	20.3	7.01	04/03/2024	11:50
5	CCV1	1	Water	NA	NA	20.3	2.01	04/03/2024	12:00
6	P1952-01	1	Solid	20.02	20	24.3	8.18	04/03/2024	12:05
7	P1952-01DUP	1	Solid	20.03	20	24.4	8.19	04/03/2024	12:06
8	P1952-04	1	Solid	20.04	20	24.2	8.14	04/03/2024	12:14
9	P1952-07	1	Solid	20.01	20	23.6	8.42	04/03/2024	12:20
10	P1952-10	1	Solid	20.03	20	25.6	8.43	04/03/2024	12:23
11	CCV2	1	Water	NA	NA	20.4	12.02	04/03/2024	12:25

WORKLIST(Hardcopy Internal Chain)

VB 130147

WorkList Name : ph p1952

WorkList ID : 179073

Department : Wet-Chemistry

Date : 04-03-2024 11:35:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1952-01	WC-A-COMP	Solid	pH	Cool 4 deg C	EART12	I31	04/02/2024	9045D
P1952-04	WC-B-COMP	Solid	pH	Cool 4 deg C	EART12	I31	04/02/2024	9045D
P1952-07	WC-C-COMP	Solid	pH	Cool 4 deg C	EART12	I31	04/02/2024	9045D
P1952-10	WC-D-COMP	Solid	pH	Cool 4 deg C	EART12	I31	04/02/2024	9045D

Date/Time 04/03/24 11:40

Raw Sample Received by: 78 Cooe

Raw Sample Relinquished by: ap sm

Date/Time 04/03/24

Raw Sample Received by: ap sm

Raw Sample Relinquished by: 78 Cooe