

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB130146

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/03/2024	11:45
2	CAL2	1	Water	NA	NA	20.3	7.00	04/03/2024	11:46
3	CAL3	1	Water	NA	NA	20.3	10.03	04/03/2024	11:47
4	ICV	1	Water	NA	NA	20.3	7.01	04/03/2024	11:50
5	CCV1	1	Water	NA	NA	20.5	2.02	04/03/2024	12:35
6	P1911-20	1	Water	NA	NA	20.8	5.72	04/03/2024	12:45
7	P1956-01	1	Water	NA	NA	25.6	7.56	04/03/2024	12:50
8	P1956-01DUP	1	Water	NA	NA	25.7	7.57	04/03/2024	12:51
9	CCV2	1	Water	NA	NA	20.3	12.03	04/03/2024	13:00

WORKLIST(Hardcopy Internal Chain)

130146

WorkList Name : ph p1956 WorkList ID : 179075 Department : Wet-Chemistry Date : 04-03-2024 11:37:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1911-20	PT-PH-WP	Water	pH	Cool 4 deg C	CHEM02	QA 01	03/26/2024	9040C
P1956-01	232	Water	pH	Cool 4 deg C	PSEG03	I31	04/02/2024	9040C

Date/Time 04/03/24 11:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04/03/24 16:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]