



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB130147

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/03/2024	11:45
2	CAL2	1	Water	NA	NA	20.3	7.00	04/03/2024	11:46
3	CAL3	1	Water	NA	NA	20.3	10.03	04/03/2024	11:47
4	ICV	1	Water	NA	NA	20.3	7.01	04/03/2024	11:50
5	CCV1	1	Water	NA	NA	20.5	2.02	04/03/2024	12:35
6	P1911-21	1	Water	NA	NA	20.8	5.72	04/03/2024	12:45
7	CCV2	1	Water	NA	NA	20.3	12.03	04/03/2024	13:00

VB 130147

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1911-1 WorkList ID : 179076 Department : Wet-Chemistry Date : 04-03-2024 11:39:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1911-21	PT-PH-WP	Water	pH	Cool 4 deg C	CHEM02	QA Of	03/26/2024	SM 4500-H B

Date/Time 04/03/24 11:40
Raw Sample Received by: SA (QC)
Raw Sample Relinquished by: SA (QA-APIC)

Date/Time 04/03/24 16:00
Raw Sample Received by: NA
Raw Sample Relinquished by: NA