



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 4/4/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 04/03/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:12
Out Date: 04/04/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB130153

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P1952-01	WC-A-COMP	1	1.16	8.50	9.66	8.34	84.5	
P1952-03	WC-A-GRAB	2	1.16	8.63	9.79	8.46	84.6	
P1952-04	WC-B-COMP	3	1.19	8.69	9.88	8.64	85.7	
P1952-06	WC-B-GRAB	4	1.19	8.58	9.77	8.51	85.3	
P1952-07	WC-C-COMP	5	1.18	8.79	9.97	8.82	86.9	
P1952-09	WC-C-GRAB	6	1.18	8.68	9.86	8.53	84.7	
P1952-10	WC-D-COMP	7	1.12	8.73	9.85	8.11	80.1	
P1952-12	WC-D-GRAB	8	1.15	8.48	9.63	7.74	77.7	
P1974-01	SOIL-1	9	1.19	8.79	9.98	8.79	86.5	
P1974-02	SOIL-1-E2	10	1.19	8.52	9.71	8.31	83.6	
P1974-04	SOIL-2	11	1.15	8.82	9.97	8.95	88.4	
P1974-05	SOIL-2-E2	12	1.16	8.61	9.77	8.6	86.4	
P1974-07	SOIL-3	13	1.17	8.53	9.7	8.67	87.9	
P1974-08	SOIL-3-E2	14	1.12	8.47	9.59	8.43	86.3	
P1975-01	BELL-4-PIPE-SLUDGE	18	1.19	8.80	9.99	6.59	61.4	
P1975-02	BELL-4-PIPE-WIPE	19	1.00	1.00	2.00	2.00	100.0	wipe sample
P1980-01	WC1-GRAB-10-11	15	1.15	8.57	9.72	8.25	82.8	
P1980-02	WC7-GRAB-12-13	16	1.19	8.50	9.69	9.1	93.1	
P1980-03	WC13-GRAB-14-15	17	1.17	8.81	9.98	9.38	93.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$