


LB130165

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

50361



Linear

A= 0.0000e+000

B= 1.9989e-004 slope

C= -1.5452e-002 $y = mx + c$

Rho= 0.9999203

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/dl
0.00	0.000	0.061	0.061	384	0.000	384					11
0.05	0.050					442					8
0.20	0.200	0.215	0.015	1152	0.0 %	1152	0				-2
2.50	2.500	2.439	-0.061	12278	0.0 %	12278					-1
5.00	5.000	4.952	-0.048	24849	0.0 %	24849					0
7.50	7.500	7.482	-0.018	37507	0.0 %	37507					1
10.0	10.000	10.051	0.051	50361	0.0 %	50361					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	384	-		0 SFAM01.1	PPB	4/4/2024 10:11	Std
	0.2	1152	-		0.2 SFAM01.1	PPB	4/4/2024 10:13	Std
	2.5	12278	-		2.5 SFAM01.1	PPB	4/4/2024 10:15	Std
	5	24849	-		5 SFAM01.1	PPB	4/4/2024 10:18	Std
	7.5	37507	-		7.5 SFAM01.1	PPB	4/4/2024 10:20	Std
	10	50361	-		10 SFAM01.1	PPB	4/4/2024 10:22	Std
ICV078	ICV078	19572	3.8969 -		SFAM01.1	PPB	4/4/2024 10:25	SMPL
ICB078	ICB078	191	0.0227 -		SFAM01.1	PPB	4/4/2024 10:27	SMPL
CCV060	CCV060	25534	5.0886 -		SFAM01.1	PPB	4/4/2024 10:30	SMPL
CCB060	CCB060	152	0.0149 -		SFAM01.1	PPB	4/4/2024 10:32	SMPL
P1967-01	E07R9	2943	0.5728 -		SFAM01.1	PPB	4/4/2024 10:39	SMPL
P1967-02	E07S1	2903	0.5648 -		SFAM01.1	PPB	4/4/2024 10:41	SMPL
P1967-03	E07S2	2904	0.565 -		SFAM01.1	PPB	4/4/2024 10:44	SMPL
P1967-04	E07S3	1411	0.2666 -		SFAM01.1	PPB	4/4/2024 10:46	SMPL
P1967-05	E07S3D	1657	0.3158 -		SFAM01.1	PPB	4/4/2024 10:48	SMPL
P1967-06	E07S3S	14415	2.866 -		SFAM01.1	PPB	4/4/2024 10:50	SMPL
P1967-07	E07S4	881	0.1607 -		SFAM01.1	PPB	4/4/2024 10:53	SMPL
P1967-08	E07S5	1152	0.2148 -		SFAM01.1	PPB	4/4/2024 10:55	SMPL
PB159997BL	PBW997	-26	-0.0206 -		SFAM01.1	PPB	4/4/2024 10:57	SMPL
CCV061	CCV061	25772	5.1362 -		SFAM01.1	PPB	4/4/2024 11:00	SMPL
CCB061	CCB061	194	0.0233 -		SFAM01.1	PPB	4/4/2024 11:02	SMPL