

LB130165


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50361



A= 0.0000e+000

B= 1.9989e-004 slope

C= -1.5452e-002 y-intercept

Rho= 0.9999203

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.061	0.061	384	0.000	384					11
0.05	0.050					442	0				8
0.20	0.200	0.215	0.015	1152	0.0 %	1152					-2
2.50	2.500	2.439	-0.061	12278	0.0 %	12278					-1
5.00	5.000	4.952	-0.048	24849	0.0 %	24849					0
7.50	7.500	7.482	-0.018	37507	0.0 %	37507					1
10.0	10.000	10.051	0.051	50361	0.0 %	50361					

LB130175 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Apr 2024 15:23:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	196	-	0.0000	7471B	PPB	04 Apr 2024 15:26:05	S	Std
0.2 - 1	50.2	902	-	0.2000	7471B	PPB	04 Apr 2024 15:28:23	S	Std
2.5 - 1	52.5	12786	-	2.5000	7471B	PPB	04 Apr 2024 15:30:40	S	Std
5.0 - 1	55	26419	-	5.0000	7471B	PPB	04 Apr 2024 15:32:57	S	Std
7.5 - 1	57.5	40093	-	7.5000	7471B	PPB	04 Apr 2024 15:35:17	S	Std
10.0 - 1	57.0	52466	-	10.0000	7471B	PPB	04 Apr 2024 15:37:38	S	Std
ICV80 - 1	ICV80	19849	3.7657	-	7471B	PPB	04 Apr 2024 15:40:59	U	SMPL
ICB80 - 1	ICB80	-264	-0.0408	-	7471B	PPB	04 Apr 2024 15:43:17	U	SMPL
CCV66 - 1	CCV66	27072	5.1327	-	7471B	PPB	04 Apr 2024 15:45:36	U	SMPL
CCB66 - 1	CCB66	-237	-0.0357	-	7471B	PPB	04 Apr 2024 15:47:51	U	SMPL
CRA - 1	CRA	806	0.1617	-	7471B	PPB	04 Apr 2024 15:50:12	U	SMPL
HighStd - 1	HighStd	53111	10.0608	-	7471B	PPB	04 Apr 2024 15:52:27	U	SMPL
ChkStd - 1	ChkStd	37801	7.1633	-	7471B	PPB	04 Apr 2024 15:54:43	U	SMPL
PB160013BL - 1	PBS	-259	-0.0399	-	7471B	PPB	04 Apr 2024 15:57:22	U	SMPL
PB160013BS - 1	LCSS	23988	4.5491	-	7471B	PPB	04 Apr 2024 15:59:41	U	SMPL
P1952-01 - 1	WC-A-COMP	29465	5.5856	-	7471B	PPB	04 Apr 2024 16:01:57	U	SMPL
P1952-04 - 1	WC-B-COMP	16211	3.0772	-	7471B	PPB	04 Apr 2024 16:04:20	U	SMPL
P1952-07 - 1	WC-C-COMP	14847	2.8191	-	7471B	PPB	04 Apr 2024 16:06:46	U	SMPL
P1952-10 - 1	WC-D-COMP	12898	2.4502	-	7471B	PPB	04 Apr 2024 16:09:06	U	SMPL
P1954-01 - 1	RBR21147	3286	0.6311	-	7471B	PPB	04 Apr 2024 16:11:24	U	SMPL
CCV67 - 1	CCV67	25796	4.8912	-	7471B	PPB	04 Apr 2024 16:13:42	U	SMPL
CCB67 - 1	CCB67	-297	-0.0471	-	7471B	PPB	04 Apr 2024 16:15:59	U	SMPL
P1954-03 - 1	RBR21145	12702	2.4131	-	7471B	PPB	04 Apr 2024 16:18:18	U	SMPL
P1955-01 - 1	CHRT-20103	3467	0.6653	-	7471B	PPB	04 Apr 2024 16:20:34	U	SMPL
P1956-04 - 1	RBR200010	2292	0.4429	-	7471B	PPB	04 Apr 2024 16:28:13	U	SMPL
P1956-04DUP - 1	RBR200010DUP	2240	0.4331	-	7471B	PPB	04 Apr 2024 16:30:29	U	SMPL
P1956-04MS - 1	RBR200010MS	23773	4.5084	-	7471B	PPB	04 Apr 2024 16:32:45	U	SMPL
P1956-04MSD - 1	RBR200010MSD	26920	5.1040	-	7471B	PPB	04 Apr 2024 16:35:04	U	SMPL
P1957-01 - 1	NM-DRUMS-040224	3087	0.5934	-	7471B	PPB	04 Apr 2024 16:37:25	U	SMPL
P1958-01 - 1	VNJ-207	2442	0.4713	-	7471B	PPB	04 Apr 2024 16:39:45	U	SMPL
P1960-02 - 1	20071	1269	0.2493	-	7471B	PPB	04 Apr 2024 16:42:03	U	SMPL
P1974-01 - 1	SOIL-1	18046	3.4245	-	7471B	PPB	04 Apr 2024 16:44:20	U	SMPL
CCV68 - 1	CCV68	26255	4.9781	-	7471B	PPB	04 Apr 2024 16:46:35	U	SMPL
CCB68 - 1	CCB68	13	0.0116	-	7471B	PPB	04 Apr 2024 16:48:54	U	SMPL
P1974-04 - 1	SOIL-2	12348	2.3461	-	7471B	PPB	04 Apr 2024 16:51:15	U	SMPL
P1974-07 - 1	SOIL-3	12011	2.2823	-	7471B	PPB	04 Apr 2024 16:53:30	U	SMPL
P1975-01 - 1	BELL-4-PIPE-SLUDGE	5342	1.0202	-	7471B	PPB	04 Apr 2024 16:55:48	U	SMPL
P1956-04LX5 - 1		280	0.0621	-	7471B	PPB	04 Apr 2024 16:58:07	U	SMPL
P1956-04A - 1		24563	4.6579	-	7471B	PPB	04 Apr 2024 17:00:25	U	SMPL
CCV69 - 1	CCV69	26411	5.0076	-	7471B	PPB	04 Apr 2024 17:07:07	U	SMPL
CCB69 - 1	CCB69	16	0.0122	-	7471B	PPB	04 Apr 2024 17:09:23	U	SMPL