



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB130176

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	04/04/2024	13:20
2	P1956-01		1	12	100.00	>212.0	>212.0	04/04/2024	13:50
3	P1956-01DUP		1	13	100.00	>212.0	>212.0	04/04/2024	14:20
4	P1976-01		1	14	100.00	>212.0	>212.0	04/04/2024	14:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 130176

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 04/04/24 flashpoint WorkList ID : 179133 Department : Wet-Chemistry Date : 04-04-2024 12:36:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1956-01	232	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/02/2024	1010B
P1976-01	AG-OILY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/03/2024	1010B

Date/Time 04/04/2024 13:00
Raw Sample Received by: N.P. W.C
Raw Sample Relinquished by: 267 CWR 1

Date/Time 04/04/2024 15:30
Raw Sample Received by: 267 CWR 1
Raw Sample Relinquished by: N.P. W.C