



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB130180

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/04/2024	15:50
2	CAL2	1	Water	NA	NA	20.3	7.00	04/04/2024	15:51
3	CAL3	1	Water	NA	NA	20.3	10.02	04/04/2024	15:52
4	ICV	1	Water	NA	NA	20.3	7.00	04/04/2024	16:00
5	CCV1	1	Water	NA	NA	20.4	2.01	04/04/2024	16:05
6	P1976-01	1	Water	NA	NA	20.6	11.62	04/04/2024	16:20
7	P1976-01DUP	1	Water	NA	NA	20.7	11.63	04/04/2024	16:21
8	P2000-04	1	Water	NA	NA	20.8	7.28	04/04/2024	16:29
9	P2000-05	1	Water	NA	NA	20.8	4.58	04/04/2024	16:37
10	CCV2	1	Water	NA	NA	20.3	12.02	04/04/2024	16:40

VB 130180

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p2000 WorkList ID : 179144 Department : Wet-Chemistry Date : 04-04-2024 15:32:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1976-01	AG-OILY-WATER	Water	pH	Cool 4 deg C	PSEG03	I31	04/03/2024	9040C
P2000-04	60152	Water	pH	Cool 4 deg C	PSEG03	P11	04/04/2024	9040C
P2000-05	60153	Water	pH	Cool 4 deg C	PSEG03	P11	04/04/2024	9040C

Date/Time 04/04/24 15:45
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04/04/24 17:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]