

LB130194


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43515



A= 0.0000e+000

B= 2.2895e-004

C= 5.5535e-002

Rho= 0.9998002

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.086	0.086	134	0.000	134	0				7
0.2	0.200	0.214	0.014	691	0.0 %	691					-3
2.5	2.500	2.424	-0.076	10347	0.0 %	10347					-2
5.0	5.000	4.884	-0.116	21088	0.0 %	21088					1
7.5	7.500	7.574	0.074	32838	0.0 %	32838					0
10.0	10.000	10.018	0.018	43515	0.0 %	43515					

LB130194 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Apr 2024 14:27:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	134	-	0.0000	7470A	PPB	05 Apr 2024 14:33:21	S	Std
0.2 - 1	50.2	691	-	0.2000	7470A	PPB	05 Apr 2024 14:35:38	S	Std
2.5 - 1	52.5	10347	-	2.5000	7470A	PPB	05 Apr 2024 14:37:55	S	Std
5.0 - 1	55	21088	-	5.0000	7470A	PPB	05 Apr 2024 14:40:12	S	Std
7.5 - 1	57.5	32838	-	7.5000	7470A	PPB	05 Apr 2024 14:42:30	S	Std
10.0 - 1	510	43515	-	10.0000	7470A	PPB	05 Apr 2024 14:44:48	S	Std
ICV82 - 1	ICV82	16337	3.7959	-	7470A	PPB	05 Apr 2024 14:47:43	U	SMPL
ICB82 - 1	ICB82	-77	0.0379	-	7470A	PPB	05 Apr 2024 14:49:59	U	SMPL
CCV72 - 1	CCV72	21456	4.9679	-	7470A	PPB	05 Apr 2024 14:52:17	U	SMPL
CCB72 - 1	CCB72	-95	0.0338	-	7470A	PPB	05 Apr 2024 14:54:32	U	SMPL
CRA - 1	CRA	545	0.1803	-	7470A	PPB	05 Apr 2024 14:56:51	U	SMPL
HighStd - 1	HighStd	43934	10.1142	-	7470A	PPB	05 Apr 2024 14:59:06	U	SMPL
ChkStd - 1	ChkStd	30462	7.0298	-	7470A	PPB	05 Apr 2024 15:01:22	U	SMPL
PB160041BL - 1	PBW	-537	-0.0674	-	7470A	PPB	05 Apr 2024 15:03:41	U	SMPL
PB160041BS - 1	LCSW	16899	3.9245	-	7470A	PPB	05 Apr 2024 15:06:00	U	SMPL
P1952-02 - 1	WC-A-COMP	-286	-0.0099	-	7470A	PPB	05 Apr 2024 15:08:16	U	SMPL
P1952-02DUP - 1	WC-A-COMPDUP	-23	0.0503	-	7470A	PPB	05 Apr 2024 15:10:34	U	SMPL
P1952-02MS - 1	WC-A-COMPMS	17747	4.1187	-	7470A	PPB	05 Apr 2024 15:12:53	U	SMPL
P1952-02MSD - 1	WC-A-COMPMSD	18204	4.2233	-	7470A	PPB	05 Apr 2024 15:15:10	U	SMPL
P1952-05 - 1	WC-B-COMP	-256	-0.0031	-	7470A	PPB	05 Apr 2024 15:17:28	U	SMPL
CCV73 - 1	CCV73	21919	5.0739	-	7470A	PPB	05 Apr 2024 15:19:46	U	SMPL
CCB73 - 1	CCB73	-53	0.0434	-	7470A	PPB	05 Apr 2024 15:22:01	U	SMPL
P1952-08 - 1	WC-C-COMP	44	0.0656	-	7470A	PPB	05 Apr 2024 15:24:19	U	SMPL
P1952-11 - 1	WC-D-COMP	74	0.0725	-	7470A	PPB	05 Apr 2024 15:26:35	U	SMPL
P1974-03 - 1	SOIL-1	110	0.0807	-	7470A	PPB	05 Apr 2024 15:28:50	U	SMPL
P1974-06 - 1	SOIL-2	141	0.0878	-	7470A	PPB	05 Apr 2024 15:31:06	U	SMPL
P1974-09 - 1	SOIL-3	99	0.0782	-	7470A	PPB	05 Apr 2024 15:33:22	U	SMPL
PB160042BL - 1	PBW	143	0.0883	-	7470A	PPB	05 Apr 2024 15:35:38	U	SMPL
PB160042BS - 1	LCSW	16410	3.8126	-	7470A	PPB	05 Apr 2024 15:37:55	U	SMPL
P1995-02 - 1	280648	-146	0.0221	-	7470A	PPB	05 Apr 2024 15:40:12	U	SMPL
P1995-04 - 1	MOO-23-00119	21	0.0603	-	7470A	PPB	05 Apr 2024 15:42:31	U	SMPL
P1995-06 - 1	MOO-23-00113	55	0.0681	-	7470A	PPB	05 Apr 2024 15:44:46	U	SMPL
CCV74 - 1	CCV74	22099	5.1151	-	7470A	PPB	05 Apr 2024 15:47:02	U	SMPL
CCB74 - 1	CCB74	139	0.0874	-	7470A	PPB	05 Apr 2024 15:49:18	U	SMPL
P1996-03 - 1	TP1	113	0.0814	-	7470A	PPB	05 Apr 2024 15:51:36	U	SMPL
P1996-06 - 1	TP2	109	0.0805	-	7470A	PPB	05 Apr 2024 15:53:51	U	SMPL
P1996-09 - 1	TP3	151	0.0901	-	7470A	PPB	05 Apr 2024 15:56:07	U	SMPL
P1996-09DUP - 1	TP3DUP	121	0.0832	-	7470A	PPB	05 Apr 2024 15:58:23	U	SMPL
P1996-09MS - 1	TP3MS	19176	4.4459	-	7470A	PPB	05 Apr 2024 16:00:38	U	SMPL
P1996-09MSD - 1	TP3MSD	16276	3.7819	-	7470A	PPB	05 Apr 2024 16:02:55	U	SMPL
PB160044BL - 1	PBW	-191	0.0118	-	7470A	PPB	05 Apr 2024 16:05:14	U	SMPL
PB160044BS - 1	LCSW	17529	4.0688	-	7470A	PPB	05 Apr 2024 16:07:33	U	SMPL
P1976-01 - 1	AG-OILY-WATER	-133	0.0251	-	7470A	PPB	05 Apr 2024 16:09:50	U	SMPL
P1983-02 - 1	COMP	666	0.2080	-	7470A	PPB	05 Apr 2024 16:12:08	U	SMPL
CCV75 - 1	CCV75	23264	5.3818	-	7470A	PPB	05 Apr 2024 16:19:59	U	SMPL
CCB75 - 1	CCB75	112	0.0812	-	7470A	PPB	05 Apr 2024 16:22:15	U	SMPL
P1983-02DUP - 1	COMPDUP	730	0.2227	-	7470A	PPB	05 Apr 2024 16:24:34	U	SMPL
P1983-02MS - 1	COMPMS	9897	2.3214	-	7470A	PPB	05 Apr 2024 16:26:49	U	SMPL
P1983-02MSD - 1	COMPMSD	10971	2.5673	-	7470A	PPB	05 Apr 2024 16:29:05	U	SMPL
PB159972TB - 1	PB159972TB	-303	-0.0138	-	7470A	PPB	05 Apr 2024 16:31:24	U	SMPL
PB160020TB - 1	PB160020TB	46	0.0661	-	7470A	PPB	05 Apr 2024 16:33:43	U	SMPL
CCV76 - 1	CCV76	22300	5.1611	-	7470A	PPB	05 Apr 2024 16:38:39	U	SMPL
CCB76 - 1	CCB76	89	0.0759	-	7470A	PPB	05 Apr 2024 16:40:56	U	SMPL
P1952-02LX5 - 1		148	0.0894	-	7470A	PPB	05 Apr 2024 16:43:16	U	SMPL
P1952-02A - 1		16641	3.8655	-	7470A	PPB	05 Apr 2024 16:45:34	U	SMPL
P1996-09LX5 - 1		-7	0.0539	-	7470A	PPB	05 Apr 2024 16:47:50	U	SMPL
P1996-09A - 1		17014	3.9509	-	7470A	PPB	05 Apr 2024 16:50:08	U	SMPL
P1983-02LX5 - 1		-3	0.0548	-	7470A	PPB	05 Apr 2024 16:52:24	U	SMPL
P1983-02A - 1		15389	3.5788	-	7470A	PPB	05 Apr 2024 16:54:42	U	SMPL
CCV77 - 1	CCV77	22672	5.2463	-	7470A	PPB	05 Apr 2024 16:57:00	U	SMPL
CCB77 - 1	CCB77	67	0.0709	-	7470A	PPB	05 Apr 2024 17:03:59	U	SMPL