


LB130200

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4806e-004 slope

C= -1.1607e-002 y-intercept

Rho= 0.9999274

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.023	0.023	138	0.000	138					=
0.05	0.050					442	0				17
0.20	0.200	0.233	0.033	988	0.0 %	988					-3
2.50	2.500	2.433	-0.067	9853	0.0 %	9853					-1
5.00	5.000	4.961	-0.039	20046	0.0 %	20046					1
7.50	7.500	7.563	0.063	30536	0.0 %	30536					0
10.0	10.000	9.987	-0.013	40307	0.0 %	40307					

LB130200

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	138	-		0 SFAM01.1	PPB	4/5/2024 17:34	Std
	0.2 S01	988	-		0.2 SFAM01.1	PPB	4/5/2024 17:39	Std
	2.5 S02	9853	-		2.5 SFAM01.1	PPB	4/5/2024 17:42	Std
	5 S03	20046	-		5 SFAM01.1	PPB	4/5/2024 17:44	Std
	7.5 S04	30536	-		7.5 SFAM01.1	PPB	4/5/2024 17:46	Std
	10 S05	40307	-		10 SFAM01.1	PPB	4/5/2024 17:54	Std
ICV083	ICV083	16611	4.109 -		SFAM01.1	PPB	4/5/2024 17:57	SMPL
ICB083	ICB083	-151	-0.0491 -		SFAM01.1	PPB	4/5/2024 17:59	SMPL
CCV078	CCV078	20884	5.1689 -		SFAM01.1	PPB	4/5/2024 18:01	SMPL
CCB078	CCB078	-147	-0.0481 -		SFAM01.1	PPB	4/5/2024 18:03	SMPL
PB160050BL	PBW050	131	0.0209 -		SFAM01.1	PPB	4/5/2024 18:06	SMPL
P1990-08	E07W3	110	0.0157 -		SFAM01.1	PPB	4/5/2024 18:08	SMPL
CCV079	CCV079	20620	5.1034 -		SFAM01.1	PPB	4/5/2024 18:10	SMPL
CCB079	CCB079	-144	-0.0473 -		SFAM01.1	PPB	4/5/2024 18:12	SMPL