

66130207

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

4/6/2024 14:01 Reviewed by : RM Instrument ID : Konelab

Test: CNEPA-NEW

04/06/2024
RM

04/06/2024
RM

Sample Id	Result	Dil. 1 +	Response	Errors
ICV001 ICV001	95.695	0.0	0.085	
ICB001 ICB001	0.814	0.0	0.002	
CCV001 CCV001	243.944	0.0	0.215	
CCB001 CCB001	0.796	0.0	0.002	
PB160059BL PBW059	0.527	0.0	0.001	
P1940-01 MC0B68	-0.357	0.0	0.001	
P1940-02 MC0B70	0.761	0.0	0.002	
P1940-03 MC0B72	0.230	0.0	0.001	
P1940-04 MC0B74	0.056	0.0	0.001	
P1940-05 MC0B76	3.054	0.0	0.004	
P1940-06 MC0B88	-0.209	0.0	0.001	
P1940-07 MC0B90	-0.304	0.0	0.001	
P1940-08 MC0B92	-0.570	0.0	0.000	
P1940-09 MC0B94	-0.298	0.0	0.001	
P1940-10 MC0BA8	0.964	0.0	0.002	
P1940-11 MC0BA8D	1.837	0.0	0.002	
P1940-12 MC0BA8S	98.524	0.0	0.087	
PB160058BL PBS058	0.750	0.0	0.001	
P2015-01 BH5E7	1.210	0.0	0.002	
P2015-02 BH5E8	0.252	0.0	0.001	
P2015-03 BH5F1	1.805	0.0	0.002	
P2015-04 BH5E4	0.437	0.0	0.001	
P2015-05 BH5E5	3.359	0.0	0.004	
P2015-06 BH5E6	2.109	0.0	0.003	
P2015-07 BH5E9	3.501	0.0	0.004	
P2015-08 BH5F0	0.808	0.0	0.002	
P2015-09 BH5F0D	0.806	0.0	0.002	
P2015-10 BH5F0S	97.930	0.0	0.087	
P2015-11 BH5F2	1.021	0.0	0.002	
CCV002 CCV002	250.840	0.0	0.221	
CCB002 CCB002	0.847	0.0	0.002	
N	31			
Mean	26.166			
SD	65.7553			
CV%	251.30			

Aquakem v. 7.2AQ1

Results from time period:

Sat Apr 06 11:01:31 2024

Sat Apr 06 13:32:17 2024

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time
S0.0	A	CNEPA-NEW	P	1.4343	µg/l	4/6/2024 11:57:59
S5.0	A	CNEPA-NEW	P	3.8797	µg/l	4/6/2024 11:58:00
S10.0	A	CNEPA-NEW	P	8.2357	µg/l	4/6/2024 11:58:01
S100.0	A	CNEPA-NEW	P	100.28	µg/l	4/6/2024 11:58:02
S250.0	A	CNEPA-NEW	P	252.3596	µg/l	4/6/2024 11:58:03
S500.0	A	CNEPA-NEW	P	498.8107	µg/l	4/6/2024 11:58:04
ICV001 ICV001	S	CNEPA-NEW	P	95.6951	µg/l	4/6/2024 13:02:49
ICB001 ICB001	S	CNEPA-NEW	P	0.8138	µg/l	4/6/2024 13:02:50
CCV001 CCV001	S	CNEPA-NEW	P	243.9437	µg/l	4/6/2024 13:02:52
CCB001 CCB001	S	CNEPA-NEW	P	0.7958	µg/l	4/6/2024 13:02:54
PB160059BL PBW059	S	CNEPA-NEW	P	0.5273	µg/l	4/6/2024 13:02:56
P1940-01 MCOB68	S	CNEPA-NEW	P	-0.3574	µg/l	4/6/2024 13:02:58
P1940-02 MCOB70	S	CNEPA-NEW	P	0.7613	µg/l	4/6/2024 13:10:23
P1940-03 MCOB72	S	CNEPA-NEW	P	0.2303	µg/l	4/6/2024 13:10:24
P1940-04 MCOB74	S	CNEPA-NEW	P	0.0561	µg/l	4/6/2024 13:10:25
P1940-05 MCOB76	S	CNEPA-NEW	P	3.0539	µg/l	4/6/2024 13:10:26
P1940-06 MCOB88	S	CNEPA-NEW	P	-0.209	µg/l	4/6/2024 13:10:27
P1940-07 MCOB90	S	CNEPA-NEW	P	-0.3036	µg/l	4/6/2024 13:10:28
P1940-08 MCOB92	S	CNEPA-NEW	P	-0.5705	µg/l	4/6/2024 13:10:29
P1940-09 MCOB94	S	CNEPA-NEW	P	-0.2979	µg/l	4/6/2024 13:10:30
P1940-10 MCOBA8	S	CNEPA-NEW	P	0.9643	µg/l	4/6/2024 13:10:31
P1940-11 MCOBA8D	S	CNEPA-NEW	P	1.8372	µg/l	4/6/2024 13:10:33
P1940-12 MCOBA8S	S	CNEPA-NEW	P	98.524	µg/l	4/6/2024 13:17:59
PB160058BL PBS058	S	CNEPA-NEW	P	0.7498	µg/l	4/6/2024 13:18:02
P2015-01 BH5E7	S	CNEPA-NEW	P	1.2097	µg/l	4/6/2024 13:18:04
P2015-02 BH5E8	S	CNEPA-NEW	P	0.2523	µg/l	4/6/2024 13:18:05
P2015-03 BH5F1	S	CNEPA-NEW	P	1.8048	µg/l	4/6/2024 13:18:06
P2015-04 BH5E4	S	CNEPA-NEW	P	0.4373	µg/l	4/6/2024 13:18:07
P2015-05 BH5E5	S	CNEPA-NEW	P	3.3593	µg/l	4/6/2024 13:18:08
P2015-06 BH5E6	S	CNEPA-NEW	P	2.1088	µg/l	4/6/2024 13:25:30
P2015-07 BH5E9	S	CNEPA-NEW	P	3.5014	µg/l	4/6/2024 13:25:31
P2015-08 BH5F0	S	CNEPA-NEW	P	0.8077	µg/l	4/6/2024 13:25:32
P2015-09 BH5F0D	S	CNEPA-NEW	P	0.8059	µg/l	4/6/2024 13:25:35
P2015-10 BH5F0S	S	CNEPA-NEW	P	97.9297	µg/l	4/6/2024 13:25:38
P2015-11 BH5F2	S	CNEPA-NEW	P	1.0209	µg/l	4/6/2024 13:25:40
CCV002 CCV002	S	CNEPA-NEW	P	250.8398	µg/l	4/6/2024 13:32:14
CCB002 CCB002	S	CNEPA-NEW	P	0.8471	µg/l	4/6/2024 13:32:16

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

4/6/2024 12:07

Test CNEPA-NEW

Accepted


4/6/2024 12:07

~~Factor~~ slope
Bias y-intercept1139- 0.000878
0.001 04/8/2024
RM

Coeff. of det.

0.999932

Errors



	Calibrator	Response	Calc. con.	Conc.	Rel. Errors
1	0.0PPBCN 50.0	0.002	1.4343	0.0000	-
2	5.0PPBCN 5.0	0.004	3.8797	5.0000	-22.4
3	10PPBCN 10.0	0.008	8.2357	10.0000	-17.6
4	100PPBCN 100.0	0.089	100.2800	100.0000	0.3
5	250PPBCN 250.0	0.222	252.3596	250.0000	0.9
6	500PPBCN 500.0	0.439	498.8107	500.0000	-0.2

04/06/2024
RM