


LB130211

SFAM01.1 INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/011
0.00	0.000	-0.046	-0.046	13	0.000	13					=
0.05	0.050					442	0				-4
0.20	0.200	0.193	-0.007	1040	0.0 %	1040					-3
2.50	2.500	2.432	-0.068	10669	0.0 %	10669					2
5.00	5.000	5.106	0.106	22166	0.0 %	22166					3
7.50	7.500	7.762	0.262	33584	0.0 %	33584					-2
10.0	10.000	9.753	-0.247	42146	0.0 %	42146					

LB130211

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	13	-		0 SFAM01.1	PPB	4/7/2024 10:55	Std
	0.2 S01	1040	-		0.2 SFAM01.1	PPB	4/7/2024 10:58	Std
	2.5 S02	10669	-		2.5 SFAM01.1	PPB	4/7/2024 11:01	Std
	5 S03	22166	-		5 SFAM01.1	PPB	4/7/2024 11:06	Std
	7.5 S04	33584	-		7.5 SFAM01.1	PPB	4/7/2024 11:08	Std
	10 S05	42146	-		10 SFAM01.1	PPB	4/7/2024 11:10	Std
ICV084	ICV084	17702	4.068 -		SFAM01.1	PPB	4/7/2024 11:13	SMPL
ICB084	ICB084	-202	-0.096 -		SFAM01.1	PPB	4/7/2024 11:15	SMPL
CCV080	CCV080	22772	5.2471 -		SFAM01.1	PPB	4/7/2024 11:18	SMPL
CCB080	CCB080	-200	-0.0955 -		SFAM01.1	PPB	4/7/2024 11:20	SMPL
PB160085BL	PBW085	89	-0.0283 -		SFAM01.1	PPB	4/7/2024 11:22	SMPL
PB160046TB	LEB046	75	-0.0316 -		SFAM01.1	PPB	4/7/2024 11:24	SMPL
P2016-01	BH5E7	261	0.0117 -		SFAM01.1	PPB	4/7/2024 11:27	SMPL
P2016-02	BH5E8	403	0.0447 -		SFAM01.1	PPB	4/7/2024 11:29	SMPL
P2016-03	BH5F1	456	0.057 -		SFAM01.1	PPB	4/7/2024 11:31	SMPL
P2016-04	BH5E4	5649	1.2648 -		SFAM01.1	PPB	4/7/2024 11:33	SMPL
P2016-05	BH5E5	3316	0.7222 -		SFAM01.1	PPB	4/7/2024 11:36	SMPL
P2016-06	BH5E6	150	-0.0141 -		SFAM01.1	PPB	4/7/2024 11:38	SMPL
P2016-07	BH5E9	13862	3.1749 -		SFAM01.1	PPB	4/7/2024 11:40	SMPL
P2016-08	BH5F0	-21	-0.0539 -		SFAM01.1	PPB	4/7/2024 11:43	SMPL
P2016-09	BH5F0D	147	-0.0148 -		SFAM01.1	PPB	4/7/2024 11:45	SMPL
P2016-10	BH5F0S	4698	1.0436 -		SFAM01.1	PPB	4/7/2024 11:50	SMPL
P2016-11	BH5F2	133	-0.0181 -		SFAM01.1	PPB	4/7/2024 11:52	SMPL
CCV081	CCV081	21270	4.8978 -		SFAM01.1	PPB	4/7/2024 11:54	SMPL
CCB081	CCB081	-67	-0.0646 -		SFAM01.1	PPB	4/7/2024 11:57	SMPL