


LB130212

SFAM01.1 / INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.1443e-004 slope

C= -4.2834e-002 Intercept

Rho= 0.9997927

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0
0.00	0.000	0.027	0.027	327	0.000	327					=
0.05	0.050					442	0				4
0.20	0.200	0.207	0.007	1165	0.0 %	1165					-2
2.50	2.500	2.438	-0.062	11571	0.0 %	11571					-1
5.00	5.000	4.958	-0.042	23320	0.0 %	23320					2
7.50	7.500	7.647	0.147	35861	0.0 %	35861					-1
10.0	10.000	9.923	-0.077	46476	0.0 %	46476					

LB130212

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	327	-		0 SFAM01.1	PPB	4/7/2024 14:53	Std
	0.2 S01	1165	-		0.2 SFAM01.1	PPB	4/7/2024 14:55	Std
	2.5 S02	11571	-		2.5 SFAM01.1	PPB	4/7/2024 14:57	Std
	5 S03	23320	-		5 SFAM01.1	PPB	4/7/2024 15:00	Std
	7.5 S04	35861	-		7.5 SFAM01.1	PPB	4/7/2024 15:05	Std
	10 S05	46476	-		10 SFAM01.1	PPB	4/7/2024 15:07	Std
ICV085	ICV085	17964	3.8092 -		SFAM01.1	PPB	4/7/2024 15:10	SMPL
ICB085	ICB085	-10	-0.045 -		SFAM01.1	PPB	4/7/2024 15:12	SMPL
CCV082	CCV082	23102	4.9109 -		SFAM01.1	PPB	4/7/2024 15:14	SMPL
CCB082	CCB082	-28	-0.0488 -		SFAM01.1	PPB	4/7/2024 15:17	SMPL
PB160090BL	PBS090	96	-0.0222 -		SFAM01.1	PPB	4/7/2024 15:19	SMPL
P2015-01	BH5E7	5665	1.1719 -		SFAM01.1	PPB	4/7/2024 15:21	SMPL
P2015-02	BH5E8	4518	0.926 -		SFAM01.1	PPB	4/7/2024 15:23	SMPL
P2015-03	BH5F1	231	0.0067 -		SFAM01.1	PPB	4/7/2024 15:26	SMPL
P2015-04	BH5E4	1317	0.2396 -		SFAM01.1	PPB	4/7/2024 15:28	SMPL
P2015-05	BH5E5	1510	0.281 -		SFAM01.1	PPB	4/7/2024 15:30	SMPL
P2015-06	BH5E6	1134	0.2003 -		SFAM01.1	PPB	4/7/2024 15:33	SMPL
P2015-07	BH5E9	5264	1.0859 -		SFAM01.1	PPB	4/7/2024 15:35	SMPL
P2015-08	BH5F0	255	0.0118 -		SFAM01.1	PPB	4/7/2024 15:37	SMPL
P2015-09	BH5F0D	320	0.0258 -		SFAM01.1	PPB	4/7/2024 15:39	SMPL
P2015-10	BH5F0S	11637	2.4525 -		SFAM01.1	PPB	4/7/2024 15:42	SMPL
P2015-11	BH5F2	262	0.0133 -		SFAM01.1	PPB	4/7/2024 15:44	SMPL
CCV083	CCV083	21315	4.5277 -		SFAM01.1	PPB	4/7/2024 15:49	SMPL
CCB083	CCB083	-75	-0.0589 -		SFAM01.1	PPB	4/7/2024 15:51	SMPL