

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Rubina

Run Number: LB130215

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/08/2024	09:10
2	CAL2	1	Water	NA	NA	20.2	7.00	04/08/2024	09:11
3	CAL3	1	Water	NA	NA	20.2	10.02	04/08/2024	09:11
4	ICV	1	Water	NA	NA	20.3	7.01	04/08/2024	09:15
5	CCV1	1	Water	NA	NA	20.3	2.01	04/08/2024	09:20
6	P1981-01	1	Solid	20.02	20	24.2	8.07	04/08/2024	09:25
7	P1981-02	1	Solid	20.02	20	24.2	8.07	04/08/2024	09:25
8	P1981-03	1	Solid	20.02	20	24.2	8.07	04/08/2024	09:25
9	CCV2	1	Water	NA	NA	20.3	12.02	04/08/2024	09:30

WORKLIST(Hardcopy Internal Chain)

VB 130215

WorkList Name : ph p1981 WorkList ID : 179211 Department : Wet-Chemistry Date : 04-08-2024 08:54:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1981-01	EOM18	Solid	pH	Cool 4 deg C	USEP01	A11	04/02/2024	9045D
P1981-02	EOM18D	Solid	pH	Cool 4 deg C	USEP01	A11	04/02/2024	9045D
P1981-03	EOM18S	Solid	pH	Cool 4 deg C	USEP01	A11	04/02/2024	9045D

Date/Time : 04/06/24 09:00
 Raw Sample Received by: JD (cc2C)
 Raw Sample Relinquished by: CFS

Date/Time : 04/06/24 12:00
 Raw Sample Received by: CFS
 Raw Sample Relinquished by: JD (cc2C)