

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Rubina

Run Number: LB130216

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/08/2024	09:10
2	CAL2	1	Water	NA	NA	20.2	7.00	04/08/2024	09:11
3	CAL3	1	Water	NA	NA	20.2	10.02	04/08/2024	09:11
4	ICV	1	Water	NA	NA	20.3	7.01	04/08/2024	09:15
5	CCV1	1	Water	NA	NA	20.3	2.01	04/08/2024	09:20
6	P2010-01	1	Water	NA	NA	20.2	7.36	04/08/2024	09:40
7	P2019-01	1	Water	NA	NA	21.2	6.51	04/08/2024	09:45
8	P2019-01DUP	1	Water	NA	NA	21.3	6.52	04/08/2024	09:46
9	CCV2	1	Water	NA	NA	20.3	12.02	04/08/2024	09:50

VB130216

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P2010 WorkList ID : 179192 Department : Wet-Chemistry Date : 04-06-2024 08:23:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2010-01	E0LZ9	Water	pH	Cool 4 deg C	USEP04	A11	04/03/2024	9040C
P2019-01	NB205-WATER	Water	pH	Cool 4 deg C	PSEG03	I41	04/05/2024	9040C

Date/Time 04/06/24 08:23:52
Raw Sample Received by: JDC (WC)
Raw Sample Relinquished by: JDC

Date/Time 04/06/24 12:30
Raw Sample Received by: JDC
Raw Sample Relinquished by: JDC (WC)