

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130229

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	04/09/2024	12:30
2	P2026-01		1	12	100.00	>212.0	>212.0	04/09/2024	13:00
3	P2026-01DUP		1	14	100.00	>212.0	>212.0	04/09/2024	13:30
4	P2026-02		1	13	100.00	>212.0	>212.0	04/09/2024	14:00
5	P2026-03		1	12	100.00	>212.0	>212.0	04/09/2024	14:30
6	P2026-04		1	13	100.00	>212.0	>212.0	04/09/2024	15:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-130229

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 04/09/24 flashpoint

WorkList ID : 179230

Department : Wet-Chemistry

Date : 04-09-2024 08:00:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2026-01	50308	Water	Flash Point	Cool 4 deg C	PSEG03	P11	04/08/2024	1010B
P2026-02	50306	Water	Flash Point	Cool 4 deg C	PSEG03	P11	04/08/2024	1010B
P2026-03	50312	Water	Flash Point	Cool 4 deg C	PSEG03	P11	04/08/2024	1010B
P2026-04	50313	Water	Flash Point	Cool 4 deg C	PSEG03	P11	04/08/2024	1010B

DateTime 04/09/2024 12:20

Raw Sample Received by: A.P.W.C

Raw Sample Relinquished by: A.W.C

Date/Time 04/09/2024 15:45

Raw Sample Received by: A.W.C

Raw Sample Relinquished by: A.P.W.C