


LB130231

SFAM01.1

/ INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.00	0.000	0.043	0.043	387	0.000	387					
0.05	0.050					442					
0.20	0.200	0.238	0.038	1262	0.0 %	1262	0				
2.50	2.500	2.481	-0.019	11279	0.0 %	11279					
5.00	5.000	4.832	-0.168	21780	0.0 %	21780					
7.50	7.500	7.586	0.086	34080	0.0 %	34080					
10.0	10.000	10.019	0.019	44944	0.0 %	44944					

0/0/0

19

1

3

1

0

LB130231

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	387	-		0 SFAM01.1	PPB	4/9/2024 9:44	Std
	0.2 S01	1262	-		0.2 SFAM01.1	PPB	4/9/2024 9:47	Std
	2.5 S02	11279	-		2.5 SFAM01.1	PPB	4/9/2024 9:49	Std
	5 S03	21780	-		5 SFAM01.1	PPB	4/9/2024 9:51	Std
	7.5 S04	34080	-		7.5 SFAM01.1	PPB	4/9/2024 9:55	Std
	10 S05	44944	-		10 SFAM01.1	PPB	4/9/2024 9:58	Std
ICV088	ICV088	17084	3.781 -		SFAM01.1	PPB	4/9/2024 10:00	SMPL
ICB088	ICB088	44	-0.0343 -		SFAM01.1	PPB	4/9/2024 10:03	SMPL
CCV092	CCV092	22394	4.9699 -		SFAM01.1	PPB	4/9/2024 10:05	SMPL
CCB092	CCB092	-13	-0.047 -		SFAM01.1	PPB	4/9/2024 10:07	SMPL
PB160040BL	PBW040	55	-0.0318 -		SFAM01.1	PPB	4/9/2024 10:09	SMPL
PB159942TB	LEB942	148	-0.011 -		SFAM01.1	PPB	4/9/2024 10:12	SMPL
P1897-01	E07A3	78	-0.0267 -		SFAM01.1	PPB	4/9/2024 10:14	SMPL
P1897-02	E07D1	84	-0.0253 -		SFAM01.1	PPB	4/9/2024 10:16	SMPL
P1897-03	E07D1D	104	-0.0208 -		SFAM01.1	PPB	4/9/2024 10:18	SMPL
P1897-04	E07D1S	4745	1.0183 -		SFAM01.1	PPB	4/9/2024 10:21	SMPL
CCV093	CCV093	22066	4.8965 -		SFAM01.1	PPB	4/9/2024 10:23	SMPL
CCB093	CCB093	62	-0.0303 -		SFAM01.1	PPB	4/9/2024 10:25	SMPL