

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130234

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/09/2024	09:25
2	CAL2	1	Water	NA	NA	20.3	7.00	04/09/2024	09:26
3	CAL3	1	Water	NA	NA	20.3	10.02	04/09/2024	09:30
4	ICV	1	Water	NA	NA	20.3	7.01	04/09/2024	09:35
5	CCV1	1	Water	NA	NA	20.3	2.01	04/09/2024	09:40
6	P2025-01	1	Water	NA	NA	22.5	6.85	04/09/2024	09:45
7	P2025-03	1	Water	NA	NA	22.5	6.85	04/09/2024	09:46
8	P2026-01	1	Water	NA	NA	22.7	6.12	04/09/2024	09:50
9	P2026-02	1	Water	NA	NA	22.7	5.15	04/09/2024	09:55
10	P2026-03	1	Water	NA	NA	22.5	5.60	04/09/2024	10:00
11	P2026-04	1	Water	NA	NA	22.6	4.78	04/09/2024	10:05
12	P2026-04DUP	1	Water	NA	NA	22.7	4.79	04/09/2024	11:00
13	CCV2	1	Water	NA	NA	20.2	12.02	04/09/2024	11:05

WORKLIST(Hardcopy Internal Chain)

130234

WorkList Name : ph p2026

WorkList ID : 179235

Department : Wet-Chemistry

Date : 04-09-2024 09:11:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2025-01	P TOTE-COMP	Water	pH	Cool 4 deg C	PSEG03	P11	04/08/2024	9040C
P2025-03	P TOTE-COMP	Water	pH	Cool 4 deg C	PSEG03	P11	04/08/2024	9040C
P2026-01	C 50308	Water	pH	Cool 4 deg C	PSEG03	P11	04/08/2024	9040C
P2026-02	C 50306	Water	pH	Cool 4 deg C	PSEG03	P11	04/08/2024	9040C
P2026-03	C 50312	Water	pH	Cool 4 deg C	PSEG03	P11	04/08/2024	9040C
P2026-04	C 50313	Water	pH	Cool 4 deg C	PSEG03	P11	04/08/2024	9040C

Date/Time 04/09/24 09:20
 Raw Sample Received by: J.C. (QC)
 Raw Sample Relinquished by: J.C. (QC)

Date/Time 04/09/24 12:00
 Raw Sample Received by: J.C. (QC)
 Raw Sample Relinquished by: J.C. (QC)