


LB130.248

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.1003e-004 slope

C= -3.8785e-002 y-intercept

Rho= 0.9999766

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	r/v
0.00	0.000	0.020	0.020	278	0.000	278					
0.05	0.050					442	0				
0.20	0.200	0.202	0.002	1145	0.0 %	1145					
2.50	2.500	2.481	-0.019	11998	0.0 %	11998					
5.00	5.000	4.965	-0.035	23826	0.0 %	23826					
7.50	7.500	7.542	0.042	36095	0.0 %	36095					
10.0	10.000	9.990	-0.010	47749	0.0 %	47749					

LB130248

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	278	-		0 SFAM01.1	PPB	4/9/2024 14:34	Std
	0.2 S01	1145	-		0.2 SFAM01.1	PPB	4/9/2024 14:36	Std
	2.5 S02	11998	-		2.5 SFAM01.1	PPB	4/9/2024 14:38	Std
	5 S03	23826	-		5 SFAM01.1	PPB	4/9/2024 14:43	Std
	7.5 S04	36095	-		7.5 SFAM01.1	PPB	4/9/2024 14:45	Std
	10 S05	47749	-		10 SFAM01.1	PPB	4/9/2024 14:50	Std
ICV089	ICV089	18856	3.9215 -		SFAM01.1	PPB	4/9/2024 14:53	SMPL
ICB089	ICB089	-124	-0.0648 -		SFAM01.1	PPB	4/9/2024 14:55	SMPL
CCV094	CCV094	23522	4.9015 -		SFAM01.1	PPB	4/9/2024 14:57	SMPL
CCB094	CCB094	-140	-0.0682 -		SFAM01.1	PPB	4/9/2024 15:00	SMPL
PB160140BL	PBW140	74	-0.0232 -		SFAM01.1	PPB	4/9/2024 15:02	SMPL
PB160099TB	LEB099	102	-0.0174 -		SFAM01.1	PPB	4/9/2024 15:04	SMPL
P1969-01	E07W6	134	-0.0106 -		SFAM01.1	PPB	4/9/2024 15:06	SMPL
P1969-02	E07W6D	124	-0.0127 -		SFAM01.1	PPB	4/9/2024 15:09	SMPL
P1969-03	E07W6S	4656	0.9391 -		SFAM01.1	PPB	4/9/2024 15:11	SMPL
P1981-01	E0M18	295	0.0232 -		SFAM01.1	PPB	4/9/2024 15:13	SMPL
P1981-02	E0M18D	396	0.0444 -		SFAM01.1	PPB	4/9/2024 15:16	SMPL
P1981-03	E0M18S	5580	1.1332 -		SFAM01.1	PPB	4/9/2024 15:18	SMPL
CCV095	CCV095	21897	4.5602 -		SFAM01.1	PPB	4/9/2024 15:20	SMPL
CCB095	CCB095	-163	-0.073 -		SFAM01.1	PPB	4/9/2024 15:22	SMPL