


LB136252

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
 B= 2.0977e-004 *slope*
 C= 1.3875e-002 *-1.3875e-002*
 Rho= 0.9998430
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>1/21)</i>
0.00	0.000	0.035	0.035	99	0.000	99					=
0.05	0.050					442	0				13
0.20	0.200	0.226	0.026	1012	0.0 %	1012					-2
2.50	2.500	2.444	-0.056	11586	0.0 %	11586					1
5.00	5.000	5.025	0.025	23889	0.0 %	23889					-2
7.50	7.500	7.386	-0.114	35143	0.0 %	35143					1
10.0	10.000	10.084	0.084	48004	0.0 %	48004					

LB130252

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	99	-		0 SFAM01.1	PPB	4/9/2024 16:09	Std
	0.2 S01	1012	-		0.2 SFAM01.1	PPB	4/9/2024 16:12	Std
	2.5 S02	11586	-		2.5 SFAM01.1	PPB	4/9/2024 16:14	Std
	5 S03	23889	-		5 SFAM01.1	PPB	4/9/2024 16:19	Std
	7.5 S04	35143	-		7.5 SFAM01.1	PPB	4/9/2024 16:21	Std
	10 S05	48004	-		10 SFAM01.1	PPB	4/9/2024 16:26	Std
ICV090	ICV090	18087	3.808 -		SFAM01.1	PPB	4/9/2024 16:29	SMPL
ICB090	ICB090	-141	-0.0157 -		SFAM01.1	PPB	4/9/2024 16:31	SMPL
CCV096	CCV096	23432	4.9293 -		SFAM01.1	PPB	4/9/2024 16:33	SMPL
CCB096	CCB096	-182	-0.0243 -		SFAM01.1	PPB	4/9/2024 16:36	SMPL
PB160144BL	PBS144	119	0.0388 -		SFAM01.1	PPB	4/9/2024 16:38	SMPL
P1850-01	EOLY5	1158	0.2568 -		SFAM01.1	PPB	4/9/2024 16:52	SMPL
P1850-02	EOLY5D	1018	0.2274 -		SFAM01.1	PPB	4/9/2024 16:55	SMPL
P1850-03	EOLY5S	12661	2.6698 -		SFAM01.1	PPB	4/9/2024 16:57	SMPL
P1850-04	EOLY8	3469	0.7416 -		SFAM01.1	PPB	4/9/2024 16:59	SMPL
P1850-05	EOLY9	2931	0.6287 -		SFAM01.1	PPB	4/9/2024 17:01	SMPL
CCV097	CCV097	23305	4.9026 -		SFAM01.1	PPB	4/9/2024 17:04	SMPL
CCB097	CCB097	-97	-0.0065 -		SFAM01.1	PPB	4/9/2024 17:06	SMPL