

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130255

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	04/10/2024	09:15
2	P2025-01		1	12	100.00	>212.0	>212.0	04/10/2024	09:45
3	P2025-01DUP		1	13	100.00	>212.0	>212.0	04/10/2024	10:15
4	P2025-03		1	12	100.00	>212.0	>212.0	04/10/2024	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-13 0255

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 4/10/24 flashpoint WorkList ID : 179265 Department : Wet-Chemistry Date : 04-10-2024 08:19:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2025-01	TOTE-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	P11	04/08/2024	1010B
P2025-03	TOTE-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	P11	04/08/2024	1010B

Date/Time 04/10/2024 09:00
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: N.P.W.C.

Date/Time 04/10/2024 11:30
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: N.P.W.C.