


LB130272

SFAM01.1 / INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.1154e-004 slope

C= -5.2504e-002 y-intercept

Rho= 0.9998810

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/b)
0.00	0.000	0.010	0.010	294	0.000	294					0/0)
0.05	0.050					442					=
0.20	0.200	0.204	0.004	1211	0.0 %	1211	0				2
2.50	2.500	2.429	-0.071	11730	0.0 %	11730					-3
5.00	5.000	5.040	0.040	24073	0.0 %	24073					1
7.50	7.500	7.588	0.088	36118	0.0 %	36118					1
10.0	10.000	9.930	-0.070	47188	0.0 %	47188					1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	294	-		0 SFAM01.1	PPB	4/11/2024 11:16	Std
	0.2 S01	1211	-		0.2 SFAM01.1	PPB	4/11/2024 11:19	Std
	2.5 S02	11730	-		2.5 SFAM01.1	PPB	4/11/2024 11:21	Std
	5 S03	24073	-		5 SFAM01.1	PPB	4/11/2024 11:23	Std
	7.5 S04	36118	-		7.5 SFAM01.1	PPB	4/11/2024 11:25	Std
	10 S05	47188	-		10 SFAM01.1	PPB	4/11/2024 11:28	Std
ICV093	ICV093	19015	3.97 -		SFAM01.1	PPB	4/11/2024 11:31	SMPL
ICB093	ICB093	116	-0.028 -		SFAM01.1	PPB	4/11/2024 11:33	SMPL
CCV006	CCV006	-113	-0.0764 -		SFAM01.1	PPB	4/11/2024 11:35	SMPL
CCB006	CCB006	0	-0.0525 -		SFAM01.1	PPB	4/11/2024 11:38	SMPL
CCV006	CCV006	24449	5.1195 -		SFAM01.1	PPB	4/11/2024 11:38	SMPL
CCB006	CCB006	79	-0.0358 -		SFAM01.1	PPB	4/11/2024 11:40	SMPL
PB160177BL	PBW1787	255	0.0014 -		SFAM01.1	PPB	4/11/2024 11:43	SMPL
P1906-01	MC0BB2	220	-0.006 -		SFAM01.1	PPB	4/11/2024 11:45	SMPL
P1906-02	MC0BB4	233	-0.0032 -		SFAM01.1	PPB	4/11/2024 11:47	SMPL
P1906-03	MC0BB6	374	0.0266 -		SFAM01.1	PPB	4/11/2024 11:49	SMPL
P1906-04	MC0BB8	2095	0.3907 -		SFAM01.1	PPB	4/11/2024 11:52	SMPL
P1906-05	MC0BC0	-245	-0.1043 -		SFAM01.1	PPB	4/11/2024 11:54	SMPL
P1906-06	MC0B78	285	0.0078 -		SFAM01.1	PPB	4/11/2024 11:56	SMPL
P1906-07	MC0B80	1882	0.3456 -		SFAM01.1	PPB	4/11/2024 11:58	SMPL
P1906-08	MC0B82	98	-0.0318 -		SFAM01.1	PPB	4/11/2024 12:01	SMPL
P1906-09	MC0B84	293	0.0095 -		SFAM01.1	PPB	4/11/2024 12:03	SMPL
P1906-10	MC0B86	286	0.008 -		SFAM01.1	PPB	4/11/2024 12:05	SMPL
P1906-11	MC0BC2	291	0.0091 -		SFAM01.1	PPB	4/11/2024 12:08	SMPL
P1906-12	MC0BC4	237	-0.0024 -		SFAM01.1	PPB	4/11/2024 12:10	SMPL
P1906-13	MC0BD9	254	0.0012 -		SFAM01.1	PPB	4/11/2024 12:12	SMPL
P1906-14	MC0BE1	734	0.1028 -		SFAM01.1	PPB	4/11/2024 12:14	SMPL
P1906-15	MC0BC8	399	0.0319 -		SFAM01.1	PPB	4/11/2024 12:17	SMPL
P1906-16	MC0BC8D	418	0.0359 -		SFAM01.1	PPB	4/11/2024 12:19	SMPL
P1906-17	MC0BC8S	4632	0.9274 -		SFAM01.1	PPB	4/11/2024 12:24	SMPL
P1906-18	MC0BE3	111	-0.029 -		SFAM01.1	PPB	4/11/2024 12:26	SMPL
P1906-19	MC0BE4	215	-0.007 -		SFAM01.1	PPB	4/11/2024 12:28	SMPL
CCV007	CCV007	22401	4.6863 -		SFAM01.1	PPB	4/11/2024 12:31	SMPL
CCB007	CCB007	32	-0.0457 -		SFAM01.1	PPB	4/11/2024 12:33	SMPL
PB160178BL	PBW178	198	-0.0106 -		SFAM01.1	PPB	4/11/2024 12:35	SMPL
P1907-01	MC0BB7	319	0.015 -		SFAM01.1	PPB	4/11/2024 12:37	SMPL
P1907-02	MC0BB9	974	0.1535 -		SFAM01.1	PPB	4/11/2024 12:40	SMPL
P1907-03	MC0BC1	-16	-0.0559 -		SFAM01.1	PPB	4/11/2024 12:42	SMPL
P1907-04	MC0B79	277	0.0061 -		SFAM01.1	PPB	4/11/2024 12:44	SMPL
P1907-05	MC0B81	2461	0.4681 -		SFAM01.1	PPB	4/11/2024 12:46	SMPL
P1907-06	MC0B83	-42	-0.0614 -		SFAM01.1	PPB	4/11/2024 12:49	SMPL
P1907-07	MC0B85	239	-0.0019 -		SFAM01.1	PPB	4/11/2024 12:51	SMPL
P1907-08	MC0B87	297	0.0103 -		SFAM01.1	PPB	4/11/2024 12:53	SMPL
P1907-09	MC0BC3	290	0.0088 -		SFAM01.1	PPB	4/11/2024 12:56	SMPL
P1907-10	MC0BC5	296	0.0101 -		SFAM01.1	PPB	4/11/2024 12:58	SMPL
P1907-11	MC0BE0	416	0.0355 -		SFAM01.1	PPB	4/11/2024 13:00	SMPL

LB130272

INSTRUMENT ID : CV1

P1907-12	MC0BE2	492	0.0516 -	SFAM01.1	PPB	4/11/2024 13:02 SMPL
P1907-13	MC0BC9	441	0.0408 -	SFAM01.1	PPB	4/11/2024 13:05 SMPL
P1907-14	MC0BC9D	441	0.0408 -	SFAM01.1	PPB	4/11/2024 13:07 SMPL
P1907-15	MC0BC9S	5442	1.0987 -	SFAM01.1	PPB	4/11/2024 13:12 SMPL
P1907-16	MC0B69	246	-0.0005 -	SFAM01.1	PPB	4/11/2024 13:14 SMPL
P1907-17	MC0B71	333	0.0179 -	SFAM01.1	PPB	4/11/2024 13:16 SMPL
P1907-18	MC0B73	304	0.0118 -	SFAM01.1	PPB	4/11/2024 13:19 SMPL
P1907-19	MC0B75	311	0.0133 -	SFAM01.1	PPB	4/11/2024 13:21 SMPL
CCV008	CCV008	21734	4.5452 -	SFAM01.1	PPB	4/11/2024 13:23 SMPL
CCB008	CCB008	-142	-0.0825 -	SFAM01.1	PPB	4/11/2024 13:25 SMPL
P1907-20	MC0B77	1658	0.2982 -	SFAM01.1	PPB	4/11/2024 13:28 SMPL
P1907-21	MC0B89	9328	1.9208 -	SFAM01.1	PPB	4/11/2024 13:30 SMPL
P1907-22	MC0B91	141	-0.0227 -	SFAM01.1	PPB	4/11/2024 13:32 SMPL
CCV009	CCV009	22954	4.8032 -	SFAM01.1	PPB	4/11/2024 13:39 SMPL
CCB009	CCB009	0	-0.0525 -	SFAM01.1	PPB	4/11/2024 13:42 SMPL