

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130278

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/11/2024	11:00
2	CAL2	1	Water	NA	NA	20.3	7.00	04/11/2024	11:02
3	CAL3	1	Water	NA	NA	20.7	10.03	04/11/2024	11:05
4	ICV	1	Water	NA	NA	20.4	7.02	04/11/2024	11:10
5	CCV1	1	Water	NA	NA	20.3	2.02	04/11/2024	11:15
6	P2066-01	1	Water	NA	NA	23.9	8.39	04/11/2024	11:25
7	P2066-02	1	Water	NA	NA	23.7	8.86	04/11/2024	11:35
8	P2074-01	1	Water	NA	NA	20.2	5.00	04/11/2024	11:40
9	P2074-01DUP	1	Water	NA	NA	20.3	5.01	04/11/2024	11:41
10	CCV2	1	Water	NA	NA	20.2	12.02	04/11/2024	11:45

130278

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p2066

WorkList ID : 179303

Department : Wet-Chemistry

Date : 04-11-2024 10:42:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2066-01	T3438	Water	pH	Cool 4 deg C	PSEG03	I21	04/10/2024	9040C
P2066-02	T3484	Water	pH	Cool 4 deg C	PSEG03	I21	04/10/2024	9040C
P2074-01	JS9600	Water	pH	Cool 4 deg C	STER17	I51	03/22/2024	9040C

Date/Time 04-11-24 10:55
Raw Sample Received by: 70100C
Raw Sample Relinquished by: 70100C

Date/Time 04-11-24 15:00
Raw Sample Received by: 70100C
Raw Sample Relinquished by: 70100C