

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130282

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	04/11/2024	14:30
2	P2066-01		1	12	100.00	>212.0	>212.0	04/11/2024	15:00
3	P2066-01DUP		1	13	100.00	>212.0	>212.0	04/11/2024	15:30
4	P2066-02		1	14	100.00	>212.0	>212.0	04/11/2024	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-130282

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 04/11/24 flashpoint WorkList ID : 179326 Department : Wet-Chemistry Date : 04-11-2024 14:00:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2066-01	T3438	Water	Flash Point	Cool 4 deg C	PSEG03	I21	04/10/2024	1010B
P2066-02	T3484	Water	Flash Point	Cool 4 deg C	PSEG03	I21	04/10/2024	1010B

Date/Time 04/11/2024 02:20
Raw Sample Received by: N.P.W.C
Raw Sample Relinquished by: JH CWC

Date/Time 04/11/2024 16:30
Raw Sample Received by: JH CWC
Raw Sample Relinquished by: N.P.W.C