

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130288

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/12/2024	08:15
2	CAL2	1	Water	NA	NA	20.3	6.99	04/12/2024	08:16
3	CAL3	1	Water	NA	NA	20.4	10.03	04/12/2024	08:17
4	ICV	1	Water	NA	NA	20.4	7.00	04/12/2024	08:20
5	CCV1	1	Water	NA	NA	20.3	2.02	04/12/2024	08:44
6	P2087-01	1	Water	NA	NA	20.1	3.75	04/12/2024	08:55
7	P2087-01DUP	1	Water	NA	NA	20.2	3.76	04/12/2024	08:56
8	CCV2	1	Water	NA	NA	20.3	12.02	04/12/2024	09:00

WORKLIST(Hardcopy Internal Chain)

130288

WorkList Name : ph p2087

WorkList ID : 179330

Department : Wet-Chemistry

Date : 04-12-2024 08:01:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2087-01	60186	Water	pH	Cool 4 deg C	PSEG03	I41	04/11/2024	9040C

Date/Time 04-12-24 08:11
Raw Sample Received by: JH (WOC)
Raw Sample Relinquished by: CP SC

Date/Time 04-12-24 10:10
Raw Sample Received by: CP SC
Raw Sample Relinquished by: JH (WOC)