

LB130292


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

48047



A= 0.0000e+000

B= 2.0551e-004

C= -4.7015e-002

Rho= 0.9993828

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	C/LD
0.0	0.000	0.010	0.010	279	0.000	279					1
0.2	0.200	0.213	0.013	1264	0.0 %	1264					7
2.5	2.500	2.391	-0.109	11863	0.0 %	11863					4
5.0	5.000	5.017	0.017	24643	0.0 %	24643					0
7.5	7.500	7.742	0.242	37900	0.0 %	37900					3
10.0	10.000	9.827	-0.173	48047	0.0 %	48047					2

LB130292 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Apr 2024 10:32:59

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	279	-	0.0000	7471B	PPB	12 Apr 2024 10:40:23	S	Std
0.2 - 1	50.2	1264	-	0.2000	7471B	PPB	12 Apr 2024 10:42:40	S	Std
2.5 - 1	52.5	11863	-	2.5000	7471B	PPB	12 Apr 2024 10:44:57	S	Std
5.0 - 1	55	24643	-	5.0000	7471B	PPB	12 Apr 2024 10:47:14	S	Std
7.5 - 1	57.5	37900	-	7.5000	7471B	PPB	12 Apr 2024 10:49:33	S	Std
10.0 - 1	510	48047	-	10.0000	7471B	PPB	12 Apr 2024 10:54:42	S	Std
ICV94 - 1	ICV94	20473	4.1603	-	7471B	PPB	12 Apr 2024 10:57:45	U	SMPL
ICB94 - 1	ICB94	53	-0.0361	-	7471B	PPB	12 Apr 2024 11:00:02	U	SMPL
CCV10 - 1	CCV10	26097	5.3161	-	7471B	PPB	12 Apr 2024 11:02:22	U	SMPL
CCB10 - 1	CCB10	47	-0.0374	-	7471B	PPB	12 Apr 2024 11:04:37	U	SMPL
CRA - 1	CRA	1229	0.2056	-	7471B	PPB	12 Apr 2024 11:07:00	U	SMPL
HighStd - 1	HighStd	49409	10.1069	-	7471B	PPB	12 Apr 2024 11:09:15	U	SMPL
ChkStd - 1	ChkStd	35466	7.2415	-	7471B	PPB	12 Apr 2024 11:14:21	U	SMPL
PB160194BL - 1	PBS	473	0.0502	-	7471B	PPB	12 Apr 2024 11:16:36	U	SMPL
PB160194BS - 1	LCSS	17089	3.4649	-	7471B	PPB	12 Apr 2024 11:19:04	U	SMPL
P2022-01 - 1	SU-03-04082024	12656	2.5539	-	7471B	PPB	12 Apr 2024 11:21:21	U	SMPL
P2038-01 - 1	OK-02-040924	5734	1.1314	-	7471B	PPB	12 Apr 2024 11:23:40	U	SMPL
P2063-01 - 1	BUR-1289	2645	0.4966	-	7471B	PPB	12 Apr 2024 11:26:01	U	SMPL
P2067-01 - 1	ETGI-346	11305	2.2762	-	7471B	PPB	12 Apr 2024 11:28:21	U	SMPL
P2070-01 - 1	TP-1	6100	1.2066	-	7471B	PPB	12 Apr 2024 11:30:37	U	SMPL
CCV11 - 1	CCV11	25162	5.1240	-	7471B	PPB	12 Apr 2024 11:32:57	U	SMPL
CCB11 - 1	CCB11	136	-0.0191	-	7471B	PPB	12 Apr 2024 11:35:14	U	SMPL
P2070-03 - 1	TP-2	6659	1.3215	-	7471B	PPB	12 Apr 2024 11:37:36	U	SMPL
P2070-05 - 1	TP-3	4061	0.7876	-	7471B	PPB	12 Apr 2024 11:39:51	U	SMPL
P2070-07 - 1	TP-4	10983	2.2101	-	7471B	PPB	12 Apr 2024 11:42:09	U	SMPL
P2070-09 - 1	TP-5	1203	0.2002	-	7471B	PPB	12 Apr 2024 11:44:27	U	SMPL
P2070-09DUP - 1	TP-5DUP	1121	0.1834	-	7471B	PPB	12 Apr 2024 11:46:45	U	SMPL
P2070-09MS - 1	TP-5MS	20548	4.1758	-	7471B	PPB	12 Apr 2024 11:49:01	U	SMPL
P2070-09MSD - 1	TP-5MSD	18290	3.7117	-	7471B	PPB	12 Apr 2024 11:51:18	U	SMPL
P2071-01 - 1	PL-01-041024	1995	0.3630	-	7471B	PPB	12 Apr 2024 11:53:38	U	SMPL
P2072-01 - 1	HR-01-041024	2721	0.5122	-	7471B	PPB	12 Apr 2024 11:56:01	U	SMPL
P2072-03 - 1	HR-02-041024	2955	0.5603	-	7471B	PPB	12 Apr 2024 11:58:17	U	SMPL
CCV12 - 1	CCV12	26123	5.3215	-	7471B	PPB	12 Apr 2024 12:00:34	U	SMPL
CCB12 - 1	CCB12	173	-0.0115	-	7471B	PPB	12 Apr 2024 12:02:51	U	SMPL
P2088-01 - 1	SU-04-041124	4863	0.9524	-	7471B	PPB	12 Apr 2024 12:05:15	U	SMPL
P2090-01 - 1	TR-05-04112024	4830	0.9456	-	7471B	PPB	12 Apr 2024 12:07:31	U	SMPL
P2091-01 - 1	OR-02-041124	2155	0.3959	-	7471B	PPB	12 Apr 2024 12:09:49	U	SMPL
P2070-09LX5 - 1		119	-0.0226	-	7471B	PPB	12 Apr 2024 12:12:06	U	SMPL
P2070-09A - 1		21315	4.3334	-	7471B	PPB	12 Apr 2024 12:14:24	U	SMPL
CCV13 - 1	CCV13	25864	5.2682	-	7471B	PPB	12 Apr 2024 12:19:28	U	SMPL
CCB13 - 1	CCB13	178	-0.0104	-	7471B	PPB	12 Apr 2024 12:21:44	U	SMPL