

LB130304

SFAM01.1

INSTRUMENT ID: CY1

Linear

 μ Abs.:

43798

A= 0.0000e+000

B= 2.3022e-004

C= -2.4044e-002

Rho= 0.9998972

Accept=Accepted

Max. Conc: —

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1/2
0.00	0.000	0.025	0.025	212	0.000	212					1
0.05	0.050					442	0				16
0.20	0.200	0.232	0.032	1113	0.0 %	1113					-3
2.50	2.500	2.434	-0.066	10677	0.0 %	10677					1
5.00	5.000	5.030	0.030	21951	0.0 %	21951					-1
7.50	7.500	7.420	-0.080	32334	0.0 %	32334					1
10.0	10.000	10.059	0.059	43798	0.0 %	43798					

LB130304
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	212	-		0 SFAM01.1	PPB	4/12/2024 15:43	Std
	0.2 S01	1113	-		0.2 SFAM01.1	PPB	4/12/2024 15:45	Std
	2.5 S02	10677	-		2.5 SFAM01.1	PPB	4/12/2024 15:47	Std
	5 S03	21951	-		5 SFAM01.1	PPB	4/12/2024 15:49	Std
	7.5 S04	32334	-		7.5 SFAM01.1	PPB	4/12/2024 15:52	Std
	10 S05	43798	-		10 SFAM01.1	PPB	4/12/2024 15:54	Std
ICV095	ICV095	17603	4.0286 -		SFAM01.1	PPB	4/12/2024 15:57	SMPL
ICB095	ICB095	-45	-0.0344 -		SFAM01.1	PPB	4/12/2024 15:59	SMPL
CCV014	CCV014	22093	5.0623 -		SFAM01.1	PPB	4/12/2024 16:01	SMPL
CCB014	CCB014	-64	-0.0388 -		SFAM01.1	PPB	4/12/2024 16:04	SMPL
PB160213BL	PBW213	357	0.0581 -		SFAM01.1	PPB	4/12/2024 16:06	SMPL
P1940-01	MCOB68	891	0.1811 -		SFAM01.1	PPB	4/12/2024 16:08	SMPL
P1940-02	MCOB70	830	0.167 -		SFAM01.1	PPB	4/12/2024 16:10	SMPL
P1940-03	MCOB72	810	0.1624 -		SFAM01.1	PPB	4/12/2024 16:13	SMPL
P1940-04	MCOB74	889	0.1806 -		SFAM01.1	PPB	4/12/2024 16:15	SMPL
P1940-05	MCOB76	3168	0.7053 -		SFAM01.1	PPB	4/12/2024 16:17	SMPL
P1940-06	MCOB88	10718	2.4435 -		SFAM01.1	PPB	4/12/2024 16:19	SMPL
P1940-07	MCOB90	709	0.1392 -		SFAM01.1	PPB	4/12/2024 16:22	SMPL
P1940-08	MCOB92	1255	0.2649 -		SFAM01.1	PPB	4/12/2024 16:24	SMPL
P1940-09	MCOB94	945	0.1935 -		SFAM01.1	PPB	4/12/2024 16:26	SMPL
P1940-10	MCOBA8	1256	0.2651 -		SFAM01.1	PPB	4/12/2024 16:34	SMPL
P1940-11	MCOBA8D	1223	0.2575 -		SFAM01.1	PPB	4/12/2024 16:36	SMPL
P1940-12	MCOBA8S	3450	0.7702 -		SFAM01.1	PPB	4/12/2024 16:38	SMPL
PB160214BL	PBW214	347	0.0558 -		SFAM01.1	PPB	4/12/2024 16:40	SMPL
P1941-01	MCOB93	1040	0.2154 -		SFAM01.1	PPB	4/12/2024 16:43	SMPL
P1941-02	MCOB95	994	0.2048 -		SFAM01.1	PPB	4/12/2024 16:45	SMPL
P1941-03	MCOBA9	1125	0.235 -		SFAM01.1	PPB	4/12/2024 16:47	SMPL
P1941-04	MCOBA9D	1132	0.2366 -		SFAM01.1	PPB	4/12/2024 16:49	SMPL
P1941-05	MCOBA9S	4287	0.9629 -		SFAM01.1	PPB	4/12/2024 16:52	SMPL
CCV015	CCV015	20430	4.6794 -		SFAM01.1	PPB	4/12/2024 16:54	SMPL
CCB015	CCB015	178	0.0169 -		SFAM01.1	PPB	4/12/2024 16:56	SMPL