

**PERCENT SOLID**

Supervisor: Iwona  
 Analyst: jignesh  
 Date: 4/15/2024

OVENTEMP IN Celsius(°C): 107  
 Time IN: 17:15  
 In Date: 04/12/2024  
 Weight Check 1.0g: 1.00  
 Weight Check 10g: 10.00  
 OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103  
 Time OUT: 08:22  
 Out Date: 04/13/2024  
 Weight Check 1.0g: 1.00  
 Weight Check 10g: 10.00  
 BalanceID: M SC-4  
 Thermometer ID: % SOLID- OVEN

QC:LB130306

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2081-01	CONCRETE-JOINT-FILLER-A	26	1.00	1.00	2.00	2.00	100.0	caluk
P2081-02	CONCRETE-JOINT-FILLER-B	27	1.00	1.00	2.00	2.00	100.0	caluk
P2081-03	CONCRETE-JOINT-FILLER-C	28	1.00	1.00	2.00	2.00	100.0	caluk
P2108-03	286	24	1.19	8.73	9.92	7.29	69.9	
P2108-04	286-E2	25	1.16	8.80	9.96	6.94	65.7	
P2109-01	BC260544-1-1	1	1.00	1.00	2.00	2.00	100.0	pilc
P2109-02	BC260544-1-2	2	1.00	1.00	2.00	2.00	100.0	pilc
P2109-03	BC260544-2-1	3	1.00	1.00	2.00	2.00	100.0	pilc
P2109-04	BC260544-2-2	4	1.00	1.00	2.00	2.00	100.0	pilc
P2109-05	KMH534F-1-1	5	1.00	1.00	2.00	2.00	100.0	pilc
P2109-06	KMH534F-1-2	6	1.00	1.00	2.00	2.00	100.0	pilc
P2110-01	BASEMENT-WALL	22	1.00	1.00	2.00	2.00	100.0	wipe sample
P2110-02	BASEMENT-CEILING	23	1.00	1.00	2.00	2.00	100.0	wipe sample
P2111-01	BASEMENT-WELL	7	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
P2112-01	Y2308-0004-1-1	8	1.00	1.00	2.00	2.00	100.0	pilc
P2112-02	Y2308-0004-1-2	9	1.00	1.00	2.00	2.00	100.0	pilc
P2112-03	BC274767-1-1	10	1.00	1.00	2.00	2.00	100.0	pilc
P2112-04	BC274767-1-2	11	1.00	1.00	2.00	2.00	100.0	pilc
P2112-05	BC288026-1-1	12	1.00	1.00	2.00	2.00	100.0	pilc
P2112-06	BC288026-1-2	13	1.00	1.00	2.00	2.00	100.0	pilc
P2113-12	410	14	1.00	1.00	2.00	2.00	100.0	wipe sample
P2113-13	411	15	1.00	1.00	2.00	2.00	100.0	wipe sample
P2113-14	412	16	1.00	1.00	2.00	2.00	100.0	wipe sample
P2113-15	314	17	1.00	1.00	2.00	2.00	100.0	wipe sample
P2113-16	314B	18	1.00	1.00	2.00	2.00	100.0	wipe sample
P2113-17	314C	19	1.00	1.00	2.00	2.00	100.0	wipe sample
P2114-01	NB-08-041224	20	1.16	8.68	9.84	8.65	86.3	
P2114-02	NB-08-04122E-2	21	1.16	8.51	9.67	8.21	82.8	

**PERCENT SOLID**

Supervisor: Iwona  
Analyst: jignesh  
Date: 4/15/2024

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:15  
In Date: 04/12/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:22  
Out Date: 04/13/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID- OVEN

QC:LB130306

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2115-05	SVOC-GPC-BLANK	29	1.00	1.00	2.00	2.00	100.0	
P2115-06	PEST-GPC-BLANK	30	1.00	1.00	2.00	2.00	100.0	
P2115-07	PEST-GPC-BLANK-SPIKE	31	1.00	1.00	2.00	2.00	100.0	
P2115-08	PCB-GPC-BLANK	32	1.00	1.00	2.00	2.00	100.0	
P2115-09	PCB-GPC-BLANK-SPIKE	33	1.00	1.00	2.00	2.00	100.0	
P2115-10	SVOC-GPC2-BLANK	34	1.00	1.00	2.00	2.00	100.0	
P2115-11	PEST-GPC2-BLANK	35	1.00	1.00	2.00	2.00	100.0	
P2115-12	PEST-GPC2-BLANK-SPIKE	36	1.00	1.00	2.00	2.00	100.0	
P2115-13	PCB-GPC2-BLANK	37	1.00	1.00	2.00	2.00	100.0	
P2115-14	PCB-GPC2-BLANK-SPIKE	38	1.00	1.00	2.00	2.00	100.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$