


LB180336

SFAM01.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.028	0.028	255	0.000	255					5
0.05	0.050					442	0				7
0.20	0.200	0.214	0.014	1105	0.0 %	1105					-4
2.50	2.500	2.404	-0.096	11113	0.0 %	11113					0
5.00	5.000	5.010	0.010	23020	0.0 %	23020					2
7.50	7.500	7.613	0.113	34915	0.0 %	34915					-1
10.0	10.000	9.932	-0.068	45511	0.0 %	45511					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	255	-		0 SFAM01.1	PPB	4/16/2024 13:43	Std
	0.2 S01	1105	-		0.2 SFAM01.1	PPB	4/16/2024 13:46	Std
	2.5 S02	11113	-		2.5 SFAM01.1	PPB	4/16/2024 13:48	Std
	5 S03	23020	-		5 SFAM01.1	PPB	4/16/2024 13:50	Std
	7.5 S04	34915	-		7.5 SFAM01.1	PPB	4/16/2024 13:53	Std
	10 S05	45511	-		10 SFAM01.1	PPB	4/16/2024 13:58	Std
ICV098	ICV098	18850	4.0972 -		SFAM01.1	PPB	4/16/2024 14:00	SMPL
ICB098	ICB098	91	-0.0079 -		SFAM01.1	PPB	4/16/2024 14:03	SMPL
CCV022	CCV022	23821	5.185 -		SFAM01.1	PPB	4/16/2024 14:05	SMPL
CCB022	CCB022	69	-0.0127 -		SFAM01.1	PPB	4/16/2024 14:07	SMPL
PB160298BL	PBS298	292	0.0361 -		SFAM01.1	PPB	4/16/2024 14:10	SMPL
P1939-02	YCMQ1	3373	0.7103 -		SFAM01.1	PPB	4/16/2024 14:16	SMPL
P1939-03	YCMQ1D	3145	0.6604 -		SFAM01.1	PPB	4/16/2024 14:18	SMPL
P1939-04	YCMQ1S	15700	3.4079 -		SFAM01.1	PPB	4/16/2024 14:21	SMPL
P1939-05	YCMQ2	2349	0.4862 -		SFAM01.1	PPB	4/16/2024 14:23	SMPL
P1939-06	YCMQ3	3256	0.6847 -		SFAM01.1	PPB	4/16/2024 14:25	SMPL
P1939-07	YCMQ4	4881	1.0403 -		SFAM01.1	PPB	4/16/2024 14:28	SMPL
P1939-08	YCMQ5	3222	0.6772 -		SFAM01.1	PPB	4/16/2024 14:30	SMPL
P1939-09	YCMQ6	2626	0.5468 -		SFAM01.1	PPB	4/16/2024 14:32	SMPL
P1939-10	YCMQ7	2588	0.5385 -		SFAM01.1	PPB	4/16/2024 14:34	SMPL
P1939-11	YCMQ8	3641	0.7689 -		SFAM01.1	PPB	4/16/2024 14:37	SMPL
P1939-12	YCMQ9	2302	0.4759 -		SFAM01.1	PPB	4/16/2024 14:39	SMPL
P1939-13	YCMR0	2769	0.5781 -		SFAM01.1	PPB	4/16/2024 14:41	SMPL
P1939-14	YCMR1	3662	0.7735 -		SFAM01.1	PPB	4/16/2024 14:43	SMPL
PB160299BL	PBS299	127	0 -		SFAM01.1	PPB	4/16/2024 14:46	SMPL
P2033-01	YCMM2	2232	0.4606 -		SFAM01.1	PPB	4/16/2024 14:48	SMPL
P2033-02	YCMM7	1981	0.4057 -		SFAM01.1	PPB	4/16/2024 14:50	SMPL
P2033-03	YCMN2	4433	0.9423 -		SFAM01.1	PPB	4/16/2024 14:53	SMPL
P2033-04	YCMN4	4270	0.9066 -		SFAM01.1	PPB	4/16/2024 14:55	SMPL
P2033-05	YCMN5	1980	0.4055 -		SFAM01.1	PPB	4/16/2024 14:57	SMPL
CCV023	CCV023	25755	5.6082 -		SFAM01.1	PPB	4/16/2024 14:59	SMPL
CCB023	CCB023	140	0.0028 -		SFAM01.1	PPB	4/16/2024 15:02	SMPL
P2033-06	YCMN6	1573	0.3164 -		SFAM01.1	PPB	4/16/2024 15:04	SMPL
P2033-07	YCMN7	2229	0.4599 -		SFAM01.1	PPB	4/16/2024 15:06	SMPL
P2033-08	YCMN8	2051	0.421 -		SFAM01.1	PPB	4/16/2024 15:08	SMPL
P2033-09	YCMP9	3225	0.6779 -		SFAM01.1	PPB	4/16/2024 15:11	SMPL
P2033-10	YCMP9D	3006	0.63 -		SFAM01.1	PPB	4/16/2024 15:13	SMPL
P2033-11	YCMP9S	17765	3.8597 -		SFAM01.1	PPB	4/16/2024 15:15	SMPL
P2033-12	YCMM3	0	-0.0278 -		SFAM01.1	PPB	4/16/2024 15:17	SMPL
P2033-11	YCMP9S	14892	3.231 -		SFAM01.1	PPB	4/16/2024 15:19	SMPL
P2033-12	YCMM3	417	0.0634 -		SFAM01.1	PPB	4/16/2024 15:21	SMPL
P2033-13	YCMM4	10234	2.2117 -		SFAM01.1	PPB	4/16/2024 15:23	SMPL
P2033-14	YCMM5	5889	1.2609 -		SFAM01.1	PPB	4/16/2024 15:25	SMPL
P2033-15	YCMM8	3718	0.7858 -		SFAM01.1	PPB	4/16/2024 15:28	SMPL
P2033-16	YCMM9	6327	1.3567 -		SFAM01.1	PPB	4/16/2024 15:30	SMPL
P2033-17	YCMN0	6531	1.4014 -		SFAM01.1	PPB	4/16/2024 15:32	SMPL

LBV130336

INSTRUMENT ID : CV1

P2033-18	YCMN1	4201	0.8915 -	SFAM01.1	PPB	4/16/2024 15:35 SMPL
P2033-19	YCMP3	2891	0.6048 -	SFAM01.1	PPB	4/16/2024 15:37 SMPL
P2033-20	YCMP4	2907	0.6083 -	SFAM01.1	PPB	4/16/2024 15:39 SMPL
P2033-21	YCMP5	2120	0.4361 -	SFAM01.1	PPB	4/16/2024 15:41 SMPL
PB160300BL	PBS300	198	0.0155 -	SFAM01.1	PPB	4/16/2024 15:44 SMPL
P2034-01	YCMN9	1027	0.1969 -	SFAM01.1	PPB	4/16/2024 15:46 SMPL
P2034-02	YCMP1	1766	0.3586 -	SFAM01.1	PPB	4/16/2024 15:48 SMPL
P2034-03	YCMP1D	1705	0.3453 -	SFAM01.1	PPB	4/16/2024 15:51 SMPL
CCV024	CCV024	25765	5.6104 -	SFAM01.1	PPB	4/16/2024 15:53 SMPL
CCB024	CCB024	135	0.0017 -	SFAM01.1	PPB	4/16/2024 15:55 SMPL
P2034-04	YCMP1S	14811	3.2133 -	SFAM01.1	PPB	4/16/2024 15:57 SMPL
P2034-05	YCMP8	1685	0.3409 -	SFAM01.1	PPB	4/16/2024 16:00 SMPL
P2034-06	YCMR2	3518	0.742 -	SFAM01.1	PPB	4/16/2024 16:02 SMPL
P2034-07	YCMR3	6467	1.3874 -	SFAM01.1	PPB	4/16/2024 16:04 SMPL
P2034-08	YCMR5	27960	6.0908 -	SFAM01.1	PPB	4/16/2024 16:06 SMPL
P2034-09	YCMR7	1919	0.3921 -	SFAM01.1	PPB	4/16/2024 16:09 SMPL
P2034-10	YCTC2	53534	11.6872 -	SFAM01.1	PPB	4/16/2024 16:11 SMPL
P2034-10DLX2	YCTC2	27359	5.9592 -	SFAM01.1	PPB	4/16/2024 16:17 SMPL
CCV025	CCV025	25850	5.629 -	SFAM01.1	PPB	4/16/2024 16:19 SMPL
CCB025	CCB025	150	0.005 -	SFAM01.1	PPB	4/16/2024 16:21 SMPL