

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130340

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	04/16/2024	13:30
2	P2108-02		1	12	100.00	>212.0	>212.0	04/16/2024	14:00
3	P2108-02DUP		1	13	100.00	>212.0	>212.0	04/16/2024	14:30

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

LB-130340

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 04/16/24 flashpont WorkList ID : 179459 Department : Wet-Chemistry Date : 04-16-2024 13:06:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2108-02	T3505	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/12/2024	1010B

Date/Time 04/16/2024 13:15
 Raw Sample Received by: N.P.W.C
 Raw Sample Relinquished by: N.P.W.C

Date/Time 04/16/2024 15:30
 Raw Sample Received by: J.P.C.B.C
 Raw Sample Relinquished by: N.P.W.C