

TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 04/16/2024 17:00 TEMP1 OUT: 103 °C 04/17/2024 08:00
 TEMP2 IN: 104 °C 04/17/2024 08:30 TEMP2 OUT: 103 °C 04/17/2024 10:00
 TEMP3 IN: 540 °C 04/17/2024 10:30 TEMP3 OUT: 540 °C 04/17/2024 11:30
 TEMP4 IN: 560 °C 04/17/2024 12:00 TEMP4 OUT: 560 °C 04/17/2024 13:30

Run Number: LB130343

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB130343BL	86.9637	86.9637	100	86.9637	86.9637	86.9637	86.9637	0.0000	0
2	P2151-01	98.4729	98.4729	100	100.6440	100.644	100.1856	100.1856	0.4584	4584
3	P2151-01DUP	91.6091	91.6091	100	93.7804	93.7804	93.3216	93.3216	0.4588	4588

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

VB130343

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tvs water p2151 WorkList ID : 179466 Department : Wet-Chemistry Date : 04-16-2024 16:09:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2151-01	Q WATER-COMP	Water	TVS	Cool 4 deg C	PSEG03	I51	04/16/2024	SM2540 G

Date/Time 04-16-24 16:15
Raw Sample Received by: JDC
Raw Sample Relinquished by: JDC

Date/Time 04-16-24 17:30
Raw Sample Received by: J.E. (m)
Raw Sample Relinquished by: JDC