

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130344

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/16/2024	16:25
2	CAL2	1	Water	NA	NA	20.2	7.00	04/16/2024	16:26
3	CAL3	1	Water	NA	NA	20.3	10.02	04/16/2024	16:30
4	ICV	1	Water	NA	NA	20.2	7.00	04/16/2024	16:33
5	CCV1	1	Water	NA	NA	20.3	2.01	04/16/2024	16:37
6	P2108-02	1	Water	NA	NA	20.7	7.80	04/16/2024	16:40
7	P2108-02DUP	1	Water	NA	NA	20.8	7.81	04/16/2024	16:41
8	P2151-01	1	Water	NA	NA	20.9	7.44	04/16/2024	16:50
9	CCV2	1	Water	NA	NA	20.2	12.02	04/16/2024	16:52

WORKLIST(Hardcopy Internal Chain)

130343

WorkList Name : ph p2151

WorkList ID : 179467

Department : Wet-Chemistry

Date : 04-16-2024 16:10:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2108-02	T3505	Water	pH	Cool 4 deg C	PSEG03	I41	04/12/2024	9040C
P2151-01	WATER-COMP	Water	pH	Cool 4 deg C	PSEG03	I51	04/16/2024	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: