

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130348

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	04/17/2024	08:40
2	P2113-01		1	12	100.00	>212.0	>212.0	04/17/2024	09:10
3	P2113-01DUP		1	13	100.00	>212.0	>212.0	04/17/2024	09:40
4	P2113-06		1	14	100.00	>212.0	>212.0	04/17/2024	10:10
5	P2113-10		1	12	100.00	>212.0	>212.0	04/17/2024	10:40
6	P2113-11		1	14	100.00	>212.0	>212.0	04/17/2024	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB-130348

Worklist Name : 04/17/24 flashpoint

Worklist ID : 179479

Department : Wet-Chemistry

Date : 04-17-2024 08:18:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2113-01	COMP-1	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/12/2024	1010B
P2113-06	COMP-2	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/12/2024	1010B
P2113-10	408	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/12/2024	1010B
P2113-11	409	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/12/2024	1010B

Date/Time 04/17/2024 08:30
 Raw Sample Received by: N.P.W.C
 Raw Sample Relinquished by: J.P. W.C

Date/Time 04/17/2024 12:15
 Raw Sample Received by: J.P. W.C
 Raw Sample Relinquished by: N.P.W.C