


LB130349

SFAM01.1

/INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3531e-004 slope

C= -3.2750e-002 intercept

Rho= 0.9997338

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.033	0.033	281	0.000	281					11
0.05	0.050					442	0				9
0.20	0.200	0.217	0.017	1061	0.0 %	1061					4
2.50	2.500	2.394	-0.106	10312	0.0 %	10312					0
5.00	5.000	4.992	-0.008	21356	0.0 %	21356					2
7.50	7.500	7.652	0.152	32657	0.0 %	32657					1
10.0	10.000	9.912	-0.088	42263	0.0 %	42263					

LB130349

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	281	-		0 SFAM01.1	PPB	4/17/2024 8:57	Std
	0.2 S01	1061	-		0.2 SFAM01.1	PPB	4/17/2024 8:59	Std
	2.5 S02	10312	-		2.5 SFAM01.1	PPB	4/17/2024 9:01	Std
	5 S03	21356	-		5 SFAM01.1	PPB	4/17/2024 9:04	Std
	7.5 S04	32657	-		7.5 SFAM01.1	PPB	4/17/2024 9:06	Std
	10 S05	42263	-		10 SFAM01.1	PPB	4/17/2024 9:08	Std
ICV099	ICV099	18287	4.2703 -		SFAM01.1	PPB	4/17/2024 9:11	SMPL
ICB099	ICB099	251	0.0263 -		SFAM01.1	PPB	4/17/2024 9:13	SMPL
CCV026	CCV026	22687	5.3056 -		SFAM01.1	PPB	4/17/2024 9:16	SMPL
CCB026	CCB026	194	0.0129 -		SFAM01.1	PPB	4/17/2024 9:18	SMPL
PB160325BL	PBS325	291	0.0357 -		SFAM01.1	PPB	4/17/2024 9:20	SMPL
P2146-01	E07W9	2154	0.4741 -		SFAM01.1	PPB	4/17/2024 9:22	SMPL
P2146-02	E07X0	2367	0.5242 -		SFAM01.1	PPB	4/17/2024 9:25	SMPL
P2146-03	E07X1	614	0.1117 -		SFAM01.1	PPB	4/17/2024 9:27	SMPL
P2146-04	E07X2	423	0.0668 -		SFAM01.1	PPB	4/17/2024 9:29	SMPL
P2146-05	E07X3	222	0.0195 -		SFAM01.1	PPB	4/17/2024 9:32	SMPL
P2146-06	E07Y0	1863	0.4056 -		SFAM01.1	PPB	4/17/2024 9:34	SMPL
P2146-07	E07Y1	1238	0.2586 -		SFAM01.1	PPB	4/17/2024 9:36	SMPL
P2146-08	E07Y2	1068	0.2186 -		SFAM01.1	PPB	4/17/2024 9:38	SMPL
P2146-09	E07Y3	753	0.1444 -		SFAM01.1	PPB	4/17/2024 9:41	SMPL
P2146-10	E07Y4	533	0.0927 -		SFAM01.1	PPB	4/17/2024 9:43	SMPL
P2146-11	E07Y4D	484	0.0811 -		SFAM01.1	PPB	4/17/2024 9:45	SMPL
P2146-12	E07Y4S	12991	3.0241 -		SFAM01.1	PPB	4/17/2024 9:47	SMPL
P2146-13	E07Y5	345	0.0484 -		SFAM01.1	PPB	4/17/2024 9:50	SMPL
CCV027	CCV027	23617	5.5245 -		SFAM01.1	PPB	4/17/2024 9:52	SMPL
CCB027	CCB027	57	-0.0193 -		SFAM01.1	PPB	4/17/2024 9:54	SMPL