

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/19/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:50
In Date: 04/18/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 04/19/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB130372

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2171-01	MDCOS9	1	1.14	8.51	9.65	7.92	79.7	
P2171-02	MDCOS9D	2	1.14	8.51	9.65	7.92	79.7	
P2171-03	MDCOS9S	3	1.14	8.51	9.65	7.92	79.7	
P2171-04	MDCOT3	4	1.16	8.58	9.74	7.78	77.2	
P2171-05	MDCOT4	5	1.15	8.36	9.51	8.06	82.7	
P2171-06	MDCOT5	6	1.2	8.68	9.88	8.26	81.3	
P2171-07	MDCOT6	7	1.15	8.43	9.58	8.15	83.0	
P2171-08	MDCOT7	8	1.15	8.80	9.95	8.16	79.7	
P2171-09	MDCOT8	9	1.19	8.54	9.73	7.93	78.9	
P2171-10	MDCOT9	10	1.18	8.62	9.8	7.92	78.2	
P2171-11	MDCOU0	11	1.16	8.54	9.7	7.92	79.2	
P2171-12	MDCOY5	12	1.19	8.42	9.61	7.53	75.3	
P2171-13	MDCOY6	13	1.15	8.79	9.94	8.18	80.0	
P2171-14	MDCOZ1	14	1.12	8.69	9.81	7.57	74.2	
P2171-15	MDCOZ2	15	1.14	8.44	9.58	7.53	75.7	
P2171-16	MDCOR5	16	1.13	8.71	9.84	7.86	77.3	
P2171-17	MDCOR6	17	1.18	8.56	9.74	7.88	78.3	
P2171-18	MDCOR7	18	1.16	8.80	9.96	8.07	78.5	
P2171-19	MDCOR8	19	1.15	8.83	9.98	8.09	78.6	
P2171-20	MDCOU3	20	1.19	8.66	9.85	7.81	76.4	
P2171-21	MDCOU4	21	1.12	8.82	9.94	8.03	78.3	
P2171-22	MDCOU5	22	1.12	8.70	9.82	7.72	75.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$