

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130378

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	04/18/2024	11:45
2	P2121-01		1	12	100.00	>212.0	>212.0	04/18/2024	12:15
3	P2121-01DUP		1	13	100.00	>212.0	>212.0	04/18/2024	12:45
4	P2121-07		1	14	100.00	>212.0	>212.0	04/18/2024	13:15
5	P2151-01		1	13	100.00	>212.0	>212.0	04/18/2024	13:45

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

LB-130378

### WORKLIST(Hardcopy Internal Chain)

WorkList Name : 4/18/24 flashpoint      WorkList ID : 179523      Department : Wet-Chemistry      Date : 04-18-2024 08:10:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2121-01	1348	Water	Flash Point	Cool 4 deg C	PSEG03	I51	04/15/2024	1010B
P2121-07	4824-28	Water	Flash Point	Cool 4 deg C	PSEG03	I51	04/15/2024	1010B
P2151-01	WATER-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	I51	04/16/2024	1010B

Date/Time 4/18/2024 11:30  
 Raw Sample Received by: N.P.W.C  
 Raw Sample Relinquished by: [Signature]

Date/Time 4/18/2024 14:20  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: N.P.W.C