

LB130391

7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49111

A= 0.0000e+000

B= 2.0462e-004

C= -1.5199e-002

Rho= 0.9999762

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	195	0.000	195				
0.2	0.200	0.215	0.015	1125	0.0 %	1125				
2.5	2.500	2.474	-0.026	12167	0.0 %	12167				
5.0	5.000	4.969	-0.031	24357	0.0 %	24357				
7.5	7.500	7.483	-0.017	36647	0.0 %	36647				
10.0	10.000	10.034	0.034	49111	0.0 %	49111				

0.0
1.0
1.0
1.0
1.0
1.0
1.0
1.0
1.0
1.0
1.0

LB1303912 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Apr 2024 14:46:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	195	-	0.0000	7471B	PPB	18 Apr 2024 14:48:07	S	Std
0.2 - 1	50.2	1125	-	0.2000	7471B	PPB	18 Apr 2024 14:50:24	S	Std
2.5 - 1	52.5	12167	-	2.5000	7471B	PPB	18 Apr 2024 14:52:42	S	Std
5.0 - 1	53	24357	-	5.0000	7471B	PPB	18 Apr 2024 14:55:00	S	Std
7.5 - 1	53.5	36647	-	7.5000	7471B	PPB	18 Apr 2024 14:57:19	S	Std
10.0 - 1	510	49111	-	10.0000	7471B	PPB	18 Apr 2024 14:59:41	S	Std
ICV03 - 1	ICV03	20283	4.1351	-	7471B	PPB	18 Apr 2024 15:02:39	U	SMPL
ICB03 - 1	ICB03	-130	-0.0418	-	7471B	PPB	18 Apr 2024 15:04:56	U	SMPL
CCV35 - 1	CCV35	24460	4.9898	-	7471B	PPB	18 Apr 2024 15:07:20	U	SMPL
CCB35 - 1	CCB35	-171	-0.0502	-	7471B	PPB	18 Apr 2024 15:09:35	U	SMPL
CRA - 1	CRA	1030	0.1956	-	7471B	PPB	18 Apr 2024 15:11:55	U	SMPL
HighStd - 1	HighStd	49440	10.1011	-	7471B	PPB	18 Apr 2024 15:14:09	U	SMPL
ChkStd - 1	ChkStd	34825	7.1106	-	7471B	PPB	18 Apr 2024 15:16:25	U	SMPL
PB160377BL - 1	PBS	-368	-0.0905	-	7471B	PPB	18 Apr 2024 15:18:53	U	SMPL
PB160377BS - 1	LCSS	19151	3.9034	-	7471B	PPB	18 Apr 2024 15:24:17	U	SMPL
P2124-01 - 1	B-131-SB01	18822	3.8361	-	7471B	PPB	18 Apr 2024 15:26:33	U	SMPL
P2124-02 - 1	B-131-SB02	6862	1.3889	-	7471B	PPB	18 Apr 2024 15:28:53	U	SMPL
P2125-01 - 1	B-101-SB01	20879	4.2570	-	7471B	PPB	18 Apr 2024 15:31:16	U	SMPL
P2125-02 - 1	B-101-SB02	2036	0.4014	-	7471B	PPB	18 Apr 2024 15:33:37	U	SMPL
P2147-01 - 1	COMP-A	9886	2.0077	-	7471B	PPB	18 Apr 2024 15:35:55	U	SMPL
CCV36 - 1	CCV36	25357	5.1733	-	7471B	PPB	18 Apr 2024 15:38:12	U	SMPL
CCB36 - 1	CCB36	5	-0.0142	-	7471B	PPB	18 Apr 2024 15:40:30	U	SMPL
P2147-01DUP - 1	COMP-ADUP	10131	2.0578	-	7471B	PPB	18 Apr 2024 15:42:49	U	SMPL
P2147-01MS - 1	COMP-AMS	31765	6.4845	-	7471B	PPB	18 Apr 2024 15:45:04	U	SMPL
P2147-01MSD - 1	COMP-AMSD	33201	6.7783	-	7471B	PPB	18 Apr 2024 15:47:22	U	SMPL
P2147-02 - 1	COMP-B	6101	1.2332	-	7471B	PPB	18 Apr 2024 15:49:44	U	SMPL
P2147-03 - 1	COMP-C	16031	3.2650	-	7471B	PPB	18 Apr 2024 15:52:07	U	SMPL
P2148-01 - 1	72-11935	4388	0.8827	-	7471B	PPB	18 Apr 2024 15:54:26	U	SMPL
P2150-01 - 1	ARCOLA-SP-1	1492	0.2901	-	7471B	PPB	18 Apr 2024 15:56:46	U	SMPL
P2151-02 - 1	SOIL	4075	0.8186	-	7471B	PPB	18 Apr 2024 15:59:05	U	SMPL
P2165-01 - 1	RPI-SS-01	2768	0.5512	-	7471B	PPB	18 Apr 2024 16:01:22	U	SMPL
P2165-02 - 1	RPI-SS-02	6680	1.3516	-	7471B	PPB	18 Apr 2024 16:03:39	U	SMPL
CCV37 - 1	CCV37	26807	5.4700	-	7471B	PPB	18 Apr 2024 16:05:56	U	SMPL
CCB37 - 1	CCB37	83	0.0018	-	7471B	PPB	18 Apr 2024 16:08:14	U	SMPL
P2165-03 - 1	RPI-SS-DUP-01	3165	0.6324	-	7471B	PPB	18 Apr 2024 16:10:35	U	SMPL
P2167-01 - 1	TP-2-1	92520	18.9161	-	7471B	PPB	18 Apr 2024 16:12:50	U	SMPL
P2167-02 - 1	TP-2-2	5829	1.1775	-	7471B	PPB	18 Apr 2024 16:15:08	U	SMPL
P2147-01LX5 - 1		1624	0.3171	-	7471B	PPB	18 Apr 2024 16:17:36	U	SMPL
P2147-01A - 1		32695	6.6748	-	7471B	PPB	18 Apr 2024 16:19:54	U	SMPL
P2167-01DLX5 - 1		18467	3.7635	-	7471B	PPB	18 Apr 2024 16:27:29	U	SMPL
CCV38 - 1	CCV38	24835	5.0665	-	7471B	PPB	18 Apr 2024 16:29:45	U	SMPL
CCB38 - 1	CCB38	56	-0.0037	-	7471B	PPB	18 Apr 2024 16:32:05	U	SMPL