

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB130399

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/19/2024	08:00
2	CAL2	1	Water	NA	NA	20.2	6.99	04/19/2024	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	04/19/2024	08:05
4	ICV	1	Water	NA	NA	20.4	7.01	04/19/2024	08:10
5	CCV1	1	Water	NA	NA	20.2	2.01	04/19/2024	08:11
6	P2189-04	1	Solid	20.02	20	24.6	6.30	04/19/2024	08:15
7	P2189-04DUP	1	Solid	20.03	20	24.8	6.31	04/19/2024	08:16
8	CCV2	1	Water	NA	NA	20.3	12.02	04/19/2024	08:20

WORKLIST(Hardcopy Internal Chain)

VP 130399

WorkList Name : corrosivity p2189

WorkList ID : 179568

Department : Wet-Chemistry

Date : 04-19-2024 07:41:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2189-04	SILICA	Solid	Corrosivity	Cool 4 deg C	PSEG03	I41	04/18/2024	9045D

Date/Time 04-19-24 07:50

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 04-19-24 11:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]