

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130400

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/19/2024	08:00
2	CAL2	1	Water	NA	NA	20.2	6.99	04/19/2024	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	04/19/2024	08:05
4	ICV	1	Water	NA	NA	20.4	7.01	04/19/2024	08:10
5	CCV1	1	Water	NA	NA	20.2	2.01	04/19/2024	08:30
6	P2185-01	1	Water	NA	NA	23.0	7.52	04/19/2024	08:35
7	P2185-01DUP	1	Water	NA	NA	23.1	7.53	04/19/2024	08:36
8	P2186-03	1	Water	NA	NA	23.2	7.05	04/19/2024	08:42
9	CCV2	1	Water	NA	NA	20.3	12.02	04/19/2024	08:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p2196

WorkList ID : 179569

Department : Wet-Chemistry

Date : 04-19-2024 07:42:47

af ~~AB130299~~ *af* ~~W3130400~~

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2185-01	<i>C</i> TOTE-1	Water	pH	Cool 4 deg C	PSEG03	I51	04/18/2024	9040C
P2186-03	<i>C</i> NWB-2030	Water	pH	Cool 4 deg C	PSEG03	I41	04/18/2024	9040C

Date/Time 04-19-24 07:50
 Raw Sample Received by: 70 (gpc)
 Raw Sample Relinquished by: *af*

Date/Time 04-19-24 11:30
 Raw Sample Received by: *af*
 Raw Sample Relinquished by: *af*