

LB130404

7470A

INSTRUMENT 131CV1

Linear

 μ Abs.:

43229

A= 0.0000e+000

B= 2.3193e-004

C= 1.3838e-003

Rho= 0.9999456

Accept=Accepted

Max. Conc:—

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.051	0.051	213	0.000	213	0				0/0/0
0.2	0.200	0.213	0.013	911	0.0 %	911					7
2.5	2.500	2.428	-0.072	10463	0.0 %	10463					3
5.0	5.000	4.985	-0.015	21490	0.0 %	21490					0
7.5	7.500	7.496	-0.004	32313	0.0 %	32313					0
10.0	10.000	10.027	0.027	43229	0.0 %	43229					0

LB130404 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Apr 2024 10:30:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	213	-	0.0000	7470A	PPB	19 Apr 2024 10:50:06	S	Std
0.2 - 1	50.2	911	-	0.2000	7470A	PPB	19 Apr 2024 10:52:23	S	Std
2.5 - 1	52.5	10463	-	2.5000	7470A	PPB	19 Apr 2024 10:54:41	S	Std
5.0 - 1	55	21490	-	5.0000	7470A	PPB	19 Apr 2024 10:56:58	S	Std
7.5 - 1	57.5	32313	-	7.5000	7470A	PPB	19 Apr 2024 10:59:15	S	Std
10.0 - 1	510	43229	-	10.0000	7470A	PPB	19 Apr 2024 11:01:32	S	Std
ICV04 - 1	ICV04	16450	3.8166	-7470A	PPB	19 Apr 2024 11:04:24	U	SMPL	
ICB04 - 1	ICB04	-40	-0.0079	-7470A	PPB	19 Apr 2024 11:06:40	U	SMPL	
CCV39 - 1	CCV39	22459	5.2102	-7470A	PPB	19 Apr 2024 11:08:57	U	SMPL	
CCB39 - 1	CCB39	-31	-0.0058	-7470A	PPB	19 Apr 2024 11:11:13	U	SMPL	
CRA - 1	CRA	867	0.2025	-7470A	PPB	19 Apr 2024 11:13:30	U	SMPL	
HighStd - 1	HighStd	42922	9.9562	-7470A	PPB	19 Apr 2024 11:15:45	U	SMPL	
ChkStd - 1	ChkStd	30392	7.0501	-7470A	PPB	19 Apr 2024 11:18:01	U	SMPL	
PB160386BL - 1	PBW	-292	-0.0663	-7470A	PPB	19 Apr 2024 11:20:20	U	SMPL	
PB160386BS - 1	LCSW	17201	3.9908	-7470A	PPB	19 Apr 2024 11:22:39	U	SMPL	
P2148-03 - 1	COMP-1	-51	-0.0104	-7470A	PPB	19 Apr 2024 11:24:55	U	SMPL	
P2148-03DUP - 1	COMP-1DUP	185	0.0443	-7470A	PPB	19 Apr 2024 11:27:14	U	SMPL	
P2148-03MS - 1	COMP-1MS	18817	4.3656	-7470A	PPB	19 Apr 2024 11:29:31	U	SMPL	
P2148-03MSD - 1	COMP-1MSD	17586	4.0801	-7470A	PPB	19 Apr 2024 11:31:48	U	SMPL	
P2148-04 - 1	COMP-2	-87	-0.0188	-7470A	PPB	19 Apr 2024 11:34:05	U	SMPL	
CCV40 - 1	CCV40	22827	5.2956	-7470A	PPB	19 Apr 2024 11:36:23	U	SMPL	
CCB40 - 1	CCB40	-29	-0.0053	-7470A	PPB	19 Apr 2024 11:38:38	U	SMPL	
P2148-05 - 1	SOM-24-37	90	0.0223	-7470A	PPB	19 Apr 2024 11:40:55	U	SMPL	
P2149-01 - 1	COMP-1	207	0.0494	-7470A	PPB	19 Apr 2024 11:43:11	U	SMPL	
P2149-02 - 1	COMP-2	248	0.0589	-7470A	PPB	19 Apr 2024 11:45:27	U	SMPL	
P2150-03 - 1	ARCOLA-SP-1	257	0.0610	-7470A	PPB	19 Apr 2024 11:47:42	U	SMPL	
P2151-04 - 1	SOIL	1569	0.3653	-7470A	PPB	19 Apr 2024 11:49:59	U	SMPL	
PB160387BL - 1	PBW	-36	-0.0070	-7470A	PPB	19 Apr 2024 11:52:15	U	SMPL	
PB160387BS - 1	LCSW	19235	4.4625	-7470A	PPB	19 Apr 2024 11:54:34	U	SMPL	
P2113-10 - 1	408	4875	1.1320	-7470A	PPB	19 Apr 2024 11:56:51	U	SMPL	
P2113-11 - 1	409	4766	1.1067	-7470A	PPB	19 Apr 2024 11:59:11	U	SMPL	
P2151-01 - 1	WATER-COMP	19089	4.4286	-7470A	PPB	19 Apr 2024 12:01:30	U	SMPL	
CCV41 - 1	CCV41	23337	5.4139	-7470A	PPB	19 Apr 2024 12:03:48	U	SMPL	
CCB41 - 1	CCB41	98	0.0241	-7470A	PPB	19 Apr 2024 12:06:05	U	SMPL	
P2185-01 - 1	TOTE-1	342	0.0807	-7470A	PPB	19 Apr 2024 12:08:22	U	SMPL	
P2185-01DUP - 1	TOTE-1DUP	363	0.0856	-7470A	PPB	19 Apr 2024 12:10:38	U	SMPL	
P2185-01MS - 1	TOTE-1MS	19249	4.4657	-7470A	PPB	19 Apr 2024 12:12:53	U	SMPL	
P2185-01MSD - 1	TOTE-1MSD	19819	4.5979	-7470A	PPB	19 Apr 2024 12:15:09	U	SMPL	
PB160321TB - 1	PB160321TB	167	0.0401	-7470A	PPB	19 Apr 2024 12:17:26	U	SMPL	
P2148-03LX5 - 1		75	0.0188	-7470A	PPB	19 Apr 2024 12:19:44	U	SMPL	
P2148-03A - 1		18093	4.1976	-7470A	PPB	19 Apr 2024 12:22:00	U	SMPL	
P2185-01LX5 - 1		19	0.0058	-7470A	PPB	19 Apr 2024 12:24:17	U	SMPL	
P2185-01A - 1		19137	4.4398	-7470A	PPB	19 Apr 2024 12:26:36	U	SMPL	
CCV42 - 1	CCV42	22924	5.3181	-7470A	PPB	19 Apr 2024 12:28:52	U	SMPL	
CCB42 - 1	CCB42	1	0.0016	-7470A	PPB	19 Apr 2024 12:31:10	U	SMPL	