

LB138408

SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46202

A= 0.0000e+000

B= 2.1739e-004 slope

C= -1.6805e-002 y-intercept

Rho= 0.9999279

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.00	0.000	0.044	0.044	281	0.000	281					462
0.05	0.050					442	0				=
0.20	0.200	0.219	0.019	1083	0.0 %	1083					10
2.50	2.500	2.463	-0.037	11406	0.0 %	11406					11
5.00	5.000	4.919	-0.081	22705	0.0 %	22705					12
7.50	7.500	7.528	0.028	34708	0.0 %	34708					00
10.0	10.000	10.027	0.027	46202	0.0 %	46202					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	281	-		0 SFAM01.1	PPB	4/19/2024 14:16	Std
	0.2 S01	1083	-		0.2 SFAM01.1	PPB	4/19/2024 14:18	Std
	2.5 S02	11406	-		2.5 SFAM01.1	PPB	4/19/2024 14:20	Std
	5 S03	22705	-		5 SFAM01.1	PPB	4/19/2024 14:23	Std
	7.5 S04	34708	-		7.5 SFAM01.1	PPB	4/19/2024 14:25	Std
	10 S05	46202	-		10 SFAM01.1	PPB	4/19/2024 14:27	Std
ICV005	ICV005	18857	4.0825 -		SFAM01.1	PPB	4/19/2024 14:30	SMPL
ICB005	ICB005	16	-0.0133 -		SFAM01.1	PPB	4/19/2024 14:32	SMPL
CCV043	CCV043	22720	4.9223 -		SFAM01.1	PPB	4/19/2024 14:35	SMPL
CCB043	CCB043	10	-0.0146 -		SFAM01.1	PPB	4/19/2024 14:37	SMPL
PB160408BL	PBS408	241	0.0356 -		SFAM01.1	PPB	4/19/2024 14:39	SMPL
P2040-02	E1CY1	1248	0.2545 -		SFAM01.1	PPB	4/19/2024 14:46	SMPL
P2040-04	E1CY1S	12954	2.7993 -		SFAM01.1	PPB	4/19/2024 14:51	SMPL
P2040-05	E1CY4	4098	0.8741 -		SFAM01.1	PPB	4/19/2024 14:53	SMPL
P2040-06	E1CY0	699	0.1352 -		SFAM01.1	PPB	4/19/2024 14:55	SMPL
P2040-07	E1CY5	1634	0.3384 -		SFAM01.1	PPB	4/19/2024 14:58	SMPL
P2040-08	E1CY3	838	0.1654 -		SFAM01.1	PPB	4/19/2024 15:00	SMPL
P2040-09	E1CY6	822	0.1619 -		SFAM01.1	PPB	4/19/2024 15:02	SMPL
P2040-10	E1CX7	1236	0.2519 -		SFAM01.1	PPB	4/19/2024 15:04	SMPL
P2040-11	E1CX8	782	0.1532 -		SFAM01.1	PPB	4/19/2024 15:07	SMPL
P2040-12	E1CX9	497	0.0912 -		SFAM01.1	PPB	4/19/2024 15:09	SMPL
PB160409BL	PBS409	265	0.0408 -		SFAM01.1	PPB	4/19/2024 15:11	SMPL
P2042-01	GCGP0	1483	0.3056 -		SFAM01.1	PPB	4/19/2024 15:13	SMPL
P2042-02	GCGP0D	1464	0.3015 -		SFAM01.1	PPB	4/19/2024 15:16	SMPL
P2042-03	GCGPOS	14896	3.2214 -		SFAM01.1	PPB	4/19/2024 15:18	SMPL
P2042-04	GCGP1	1054	0.2123 -		SFAM01.1	PPB	4/19/2024 15:20	SMPL
P2042-05	GCGP2	1662	0.3445 -		SFAM01.1	PPB	4/19/2024 15:22	SMPL
P2042-06	GCGP3	1273	0.2599 -		SFAM01.1	PPB	4/19/2024 15:25	SMPL
P2042-07	GCGP4	2887	0.6108 -		SFAM01.1	PPB	4/19/2024 15:27	SMPL
CCV044	CCV044	25049	5.4286 -		SFAM01.1	PPB	4/19/2024 15:29	SMPL
CCB044	CCB044	178	0.0219 -		SFAM01.1	PPB	4/19/2024 15:32	SMPL
P2042-08	GCGP5	1154	0.2341 -		SFAM01.1	PPB	4/19/2024 15:34	SMPL
P2042-09	GCGP6	1640	0.3397 -		SFAM01.1	PPB	4/19/2024 15:36	SMPL
P2042-10	GCGP7	3080	0.6528 -		SFAM01.1	PPB	4/19/2024 15:38	SMPL
P2042-11	GCGP8	1476	0.3041 -		SFAM01.1	PPB	4/19/2024 15:41	SMPL
P2042-12	GCGP9	1496	0.3084 -		SFAM01.1	PPB	4/19/2024 15:43	SMPL
P2042-13	GCGQ0	1804	0.3754 -		SFAM01.1	PPB	4/19/2024 15:45	SMPL
P2042-14	GCGQ1	1520	0.3136 -		SFAM01.1	PPB	4/19/2024 15:47	SMPL
P2042-15	GCGQ2	2132	0.4467 -		SFAM01.1	PPB	4/19/2024 15:50	SMPL
P2042-16	GCGQ3	1572	0.3249 -		SFAM01.1	PPB	4/19/2024 15:52	SMPL
P2042-17	GCGQ4	2770	0.5854 -		SFAM01.1	PPB	4/19/2024 15:54	SMPL
P2042-18	GCGQ5	2198	0.461 -		SFAM01.1	PPB	4/19/2024 15:57	SMPL
P2042-19	GCGQ6	1375	0.2821 -		SFAM01.1	PPB	4/19/2024 15:59	SMPL
P2042-20	GCGQ7	3710	0.7897 -		SFAM01.1	PPB	4/19/2024 16:01	SMPL
P2042-21	GCGQ8	12810	2.7679 -		SFAM01.1	PPB	4/19/2024 16:03	SMPL
P2042-22	GCGQ9	15417	3.3347 -		SFAM01.1	PPB	4/19/2024 16:06	SMPL

INSTRUMENT ID : CV1

PB160410BL	PBS410	209	0.0286 -	SFAM01.1	PPB	4/19/2024 16:08 SMPL
P2043-01	GCGR0	14671	3.1725 -	SFAM01.1	PPB	4/19/2024 16:15 SMPL
P2043-02	GCGR0D	14609	3.159 -	SFAM01.1	PPB	4/19/2024 16:17 SMPL
P2043-03	GCGR0S	26958	5.8436 -	SFAM01.1	PPB	4/19/2024 16:22 SMPL
P2043-04	GCGR1	3322	0.7054 -	SFAM01.1	PPB	4/19/2024 16:25 SMPL
CCV045	CCV045	24649	5.3416 -	SFAM01.1	PPB	4/19/2024 16:27 SMPL
CCB045	CCB045	176	0.0215 -	SFAM01.1	PPB	4/19/2024 16:29 SMPL
P2043-05	GCGR2	16944	3.6666 -	SFAM01.1	PPB	4/19/2024 16:31 SMPL
P2043-06	GCGR3	2907	0.6151 -	SFAM01.1	PPB	4/19/2024 16:34 SMPL
P2043-07	GCGR4	3963	0.8447 -	SFAM01.1	PPB	4/19/2024 16:36 SMPL
P2043-08	GCGR5	4250	0.9071 -	SFAM01.1	PPB	4/19/2024 16:38 SMPL
P2043-09	GCGR6	7517	1.6173 -	SFAM01.1	PPB	4/19/2024 16:40 SMPL
P2043-10	GCGR7	1501	0.3095 -	SFAM01.1	PPB	4/19/2024 16:43 SMPL
P2043-11	GCGR8	4201	0.8964 -	SFAM01.1	PPB	4/19/2024 16:45 SMPL
P2043-12	GCGR9	2705	0.5712 -	SFAM01.1	PPB	4/19/2024 16:47 SMPL
P2043-13	GCGS0	4794	1.0254 -	SFAM01.1	PPB	4/19/2024 16:50 SMPL
P2043-14	GCGS1	3154	0.6688 -	SFAM01.1	PPB	4/19/2024 16:52 SMPL
P2043-15	GCGS2	1578	0.3262 -	SFAM01.1	PPB	4/19/2024 16:54 SMPL
P2043-16	GCGS3	3157	0.6695 -	SFAM01.1	PPB	4/19/2024 16:56 SMPL
P2043-17	GCGS4	5783	1.2404 -	SFAM01.1	PPB	4/19/2024 16:59 SMPL
P2043-18	GCGS5	952	0.1901 -	SFAM01.1	PPB	4/19/2024 17:01 SMPL
P2043-19	GCGS6	3712	0.7901 -	SFAM01.1	PPB	4/19/2024 17:03 SMPL
P2043-20	GCGS7	2609	0.5504 -	SFAM01.1	PPB	4/19/2024 17:05 SMPL
P2043-21	GCGS8	2646	0.5584 -	SFAM01.1	PPB	4/19/2024 17:08 SMPL
P2043-22	GCGS9	2324	0.4884 -	SFAM01.1	PPB	4/19/2024 17:10 SMPL
P2040-03	E1CY1D	1394	0.2862 -	SFAM01.1	PPB	4/19/2024 17:15 SMPL
CCV046	CCV046	25016	5.4214 -	SFAM01.1	PPB	4/19/2024 17:17 SMPL
CCB046	CCB046	184	0.0232 -	SFAM01.1	PPB	4/19/2024 17:19 SMPL