

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130409

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/19/2024	14:15
2	CAL2	1	Water	NA	NA	20.1	7.00	04/19/2024	14:16
3	CAL3	1	Water	NA	NA	20.3	10.02	04/19/2024	14:20
4	ICV	1	Water	NA	NA	20.3	7.02	04/19/2024	14:22
5	CCV1	1	Water	NA	NA	20.2	2.02	04/19/2024	14:25
6	P2207-01	1	Water	NA	NA	20.1	5.98	04/19/2024	14:35
7	P2207-02	1	Water	NA	NA	20.8	5.82	04/19/2024	14:40
8	P2207-02DUP	1	Water	NA	NA	20.9	5.83	04/19/2024	14:41
9	CCV2	1	Water	NA	NA	20.3	12.02	04/19/2024	14:45

WORKLIST(Hardcopy Internal Chain)

UB 130409

WorkList Name : ph p2109

WorkList ID : 179592

Department : Wet-Chemistry

Date : 04-19-2024 13:51:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2207-01	C BUR-1288	Water	pH	Cool 4 deg C	PSEG03	I41	04/19/2024	9040C
P2207-02	C BUR-1291	Water	pH	Cool 4 deg C	PSEG03	I41	04/19/2024	9040C

Date/Time 04/19/24 14:10

Raw Sample Received by: JD Cello

Raw Sample Relinquished by: AP SM

Date/Time 04/19/24 17:00

Raw Sample Received by: AP SM

Raw Sample Relinquished by: JD Cello