

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/23/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/22/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 04/23/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB130425

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2219-01	SS051P-SD12-041824-00	1	1.16	8.50	9.66	8.15	82.2	
P2219-02	SS051P-SD11-041824-00	2	1.19	8.65	9.84	7.81	76.5	
P2219-03	AT054P-SD04-041824-00	3	1.15	8.84	9.99	8.57	83.9	
P2223-01	W-1	4	1.00	1.00	2.00	2.00	100.0	wipe sample
P2223-02	W-2	5	1.00	1.00	2.00	2.00	100.0	wipe sample
P2223-03	W-3	6	1.00	1.00	2.00	2.00	100.0	wipe sample
P2223-04	W-4	7	1.00	1.00	2.00	2.00	100.0	wipe sample
P2223-05	W-5	8	1.00	1.00	2.00	2.00	100.0	wipe sample
P2225-01	TR-06-042224	9	1.13	8.46	9.59	9.22	95.6	
P2225-02	TR-06-042224-E2	10	1.12	8.45	9.57	8.71	89.8	
P2227-01	HD-02-04222024	11	1.17	8.58	9.75	9.2	93.6	
P2227-02	HD-02-04222024	12	1.12	8.55	9.67	9.02	92.4	
P2228-01	EXCAVATION-SOIL	13	1.13	8.85	9.98	8.98	88.7	
P2228-02	EXCAVATION-SOIL-E2	14	1.12	8.75	9.87	8.66	86.2	
P2232-01	PCB041824RB01	15	1.00	1.00	2.00	2.00	100.0	caluk
P2232-02	PCB041824RB02	16	1.00	1.00	2.00	2.00	100.0	caluk
P2232-03	PCB041824RB03	17	1.00	1.00	2.00	2.00	100.0	caluk
P2232-04	PCB041824RB04	18	1.00	1.00	2.00	2.00	100.0	caluk

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$