

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130428

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	04/22/2024	11:15
2	P2207-01		1	12	100.00	>212.0	>212.0	04/22/2024	11:45
3	P2207-01DUP		1	13	100.00	>212.0	>212.0	04/22/2024	12:15
4	P2207-02		1	14	100.00	>212.0	>212.0	04/22/2024	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB-130428

WorkList Name : 04/22/24 flashpoint

WorkList ID : 179629

Department : Wet-Chemistry

Date : 04-22-2024 10:35:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2207-01	BUR-1288	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/19/2024	1010B
P2207-02	BUR-1291	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/19/2024	1010B

Date/Time 04/22/24 11:00
Raw Sample Received by: N.P.-W.C.
Raw Sample Relinquished by: J.P. (COC)

Date/Time 04/22/24 13:20
Raw Sample Received by: J.P. (COC)
Raw Sample Relinquished by: N.P.-W.C.