

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130429

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	04/22/2024	13:45
2	P2185-03		1	12	100.00	>212.0	>212.0	04/22/2024	14:15
3	P2185-03DUP		1	13	100.00	>212.0	>212.0	04/22/2024	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-130429

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 4/22/24 flashpoint

WorkList ID : 179635

Department : Wet-Chemistry

Date : 04-22-2024 13:02:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2185-03	TOTE-2	Solid	Flash Point	Cool 4 deg C	PSEG03	I51	04/18/2024	1010B

Date/Time 04/22/2024 13:30

Raw Sample Received by: N.P.W.C

Raw Sample Relinquished by: JAC

Date/Time 04/22/2024 15:30

Raw Sample Received by: JAC

Raw Sample Relinquished by: N.P.W.C