

LB130436


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43944



A= 0.0000e+000

B= 2.3015e-004 slope

C= -2.8827e-002 y-intercept

Rho= 0.9998570

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	0.057	0.057	371	0.000	371					
0.05	0.050					442	0				11
0.20	0.200	0.218	0.018	1074	0.0 %	1074					9
2.50	2.500	2.487	-0.013	10930	0.0 %	10930					11
5.00	5.000	4.898	-0.102	21405	0.0 %	21405					12
7.50	7.500	7.456	-0.044	32521	0.0 %	32521					11
10.0	10.000	10.085	0.085	43944	0.0 %	43944					1

INSTRUMNT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	371	-		0 SFAM01.1	PPB	4/23/2024 8:58	Std
	0.2 S01	1074	-		0.2 SFAM01.1	PPB	4/23/2024 9:01	Std
	2.5 S02	10930	-		2.5 SFAM01.1	PPB	4/23/2024 9:03	Std
	5 S03	21405	-		5 SFAM01.1	PPB	4/23/2024 9:05	Std
	7.5 S04	32521	-		7.5 SFAM01.1	PPB	4/23/2024 9:07	Std
	10 S05	43944	-		10 SFAM01.1	PPB	4/23/2024 9:10	Std
ICV007	ICV007	18360	4.1967	-	SFAM01.1	PPB	4/23/2024 9:16	SMPL
ICB007	ICB007	217	0.0211	-	SFAM01.1	PPB	4/23/2024 9:21	SMPL
CCV051	CCV051	22018	5.0386	-	SFAM01.1	PPB	4/23/2024 9:23	SMPL
CCB051	CCB051	53	-0.0166	-	SFAM01.1	PPB	4/23/2024 9:26	SMPL
PB160437BL	PBS437	173	0.011	-	SFAM01.1	PPB	4/23/2024 9:30	SMPL
P2044-04	GCGT1	4680	1.0483	-	SFAM01.1	PPB	4/23/2024 9:39	SMPL
P2044-05	GCGT2	4583	1.026	-	SFAM01.1	PPB	4/23/2024 9:42	SMPL
P2044-06	GCGT3	2415	0.527	-	SFAM01.1	PPB	4/23/2024 9:44	SMPL
P2044-07	GCGT4	3330	0.7376	-	SFAM01.1	PPB	4/23/2024 9:46	SMPL
P2044-08	GCGT5	3387	0.7507	-	SFAM01.1	PPB	4/23/2024 9:49	SMPL
P2044-09	GCGT6	3647	0.8105	-	SFAM01.1	PPB	4/23/2024 9:51	SMPL
P2044-10	GCGT7	1905	0.4096	-	SFAM01.1	PPB	4/23/2024 9:53	SMPL
P2044-11	GCGT8	1400	0.2934	-	SFAM01.1	PPB	4/23/2024 9:55	SMPL
P2044-12	GCGT9	4402	0.9843	-	SFAM01.1	PPB	4/23/2024 9:58	SMPL
P2044-13	GCGW0	2013	0.4345	-	SFAM01.1	PPB	4/23/2024 10:00	SMPL
P2044-14	GCGW1	1864	0.4002	-	SFAM01.1	PPB	4/23/2024 10:02	SMPL
P2044-15	GCGW2	5301	1.1912	-	SFAM01.1	PPB	4/23/2024 10:04	SMPL
P2044-16	GCGW3	2098	0.454	-	SFAM01.1	PPB	4/23/2024 10:07	SMPL
P2044-17	GCGW4	2694	0.5912	-	SFAM01.1	PPB	4/23/2024 10:09	SMPL
P2044-18	GCGW5	2008	0.4333	-	SFAM01.1	PPB	4/23/2024 10:11	SMPL
P2044-19	GCGW6	1773	0.3792	-	SFAM01.1	PPB	4/23/2024 10:13	SMPL
CCV052	CCV052	23621	5.4076	-	SFAM01.1	PPB	4/23/2024 10:16	SMPL
CCB052	CCB052	-210	-0.0772	-	SFAM01.1	PPB	4/23/2024 10:18	SMPL
P2044-20	GCGW7	2110	0.4568	-	SFAM01.1	PPB	4/23/2024 10:20	SMPL
P2044-21	GCGW8	1705	0.3636	-	SFAM01.1	PPB	4/23/2024 10:23	SMPL
P2044-22	GCGW9	2892	0.6368	-	SFAM01.1	PPB	4/23/2024 10:25	SMPL
PB160438BL	PBS438	-86	-0.0486	-	SFAM01.1	PPB	4/23/2024 10:27	SMPL
P2045-01	GCGX0	2813	0.6186	-	SFAM01.1	PPB	4/23/2024 10:29	SMPL
P2045-02	GCGX1	2402	0.524	-	SFAM01.1	PPB	4/23/2024 10:32	SMPL
P2045-03	GCGX1D	2340	0.5097	-	SFAM01.1	PPB	4/23/2024 10:34	SMPL
P2045-04	GCGX1S	14488	3.3056	-	SFAM01.1	PPB	4/23/2024 10:36	SMPL
P2045-05	GCGX2	6221	1.4029	-	SFAM01.1	PPB	4/23/2024 10:38	SMPL
P2045-06	GCGX3	2916	0.6423	-	SFAM01.1	PPB	4/23/2024 10:41	SMPL
P2045-07	GCGX4	2297	0.4998	-	SFAM01.1	PPB	4/23/2024 10:43	SMPL
P2045-08	GCGX5	3461	0.7677	-	SFAM01.1	PPB	4/23/2024 10:45	SMPL
P2045-09	GCGX6	2477	0.5413	-	SFAM01.1	PPB	4/23/2024 10:48	SMPL
P2045-10	GCGX7	2831	0.6227	-	SFAM01.1	PPB	4/23/2024 10:50	SMPL
P2045-11	GCGX8	2052	0.4434	-	SFAM01.1	PPB	4/23/2024 10:52	SMPL
P2045-12	GCGX9	2264	0.4922	-	SFAM01.1	PPB	4/23/2024 10:54	SMPL
P2045-13	GCGY0	3358	0.744	-	SFAM01.1	PPB	4/23/2024 10:57	SMPL

INSTRUMNT ID : CV1

P2045-14	GCGY1	1888	0.4057 -	SFAM01.1	PPB	4/23/2024 10:59	SMPL
P2045-15	GCGY2	3592	0.7979 -	SFAM01.1	PPB	4/23/2024 11:01	SMPL
P2045-16	GCGY3	2132	0.4619 -	SFAM01.1	PPB	4/23/2024 11:03	SMPL
CCV053	CCV053	24267	5.5562 -	SFAM01.1	PPB	4/23/2024 11:06	SMPL
CCB053	CCB053	-216	-0.0785 -	SFAM01.1	PPB	4/23/2024 11:08	SMPL
P2045-17	GCGY4	3378	0.7486 -	SFAM01.1	PPB	4/23/2024 11:10	SMPL
P2045-18	GCGY5	4702	1.0533 -	SFAM01.1	PPB	4/23/2024 11:12	SMPL
P2045-19	GCGY6	1992	0.4296 -	SFAM01.1	PPB	4/23/2024 11:15	SMPL
P2045-20	GCGY7	3443	0.7636 -	SFAM01.1	PPB	4/23/2024 11:17	SMPL
P2045-21	GCGY8	2415	0.527 -	SFAM01.1	PPB	4/23/2024 11:19	SMPL
P2045-22	GCGY9	359865	HIGH -	SFAM01.1	PPB	4/23/2024 11:22	
P2044-01	GCGT0	4394	0.9825 -	SFAM01.1	PPB	4/23/2024 11:27	SMPL
P2044-02	GCGT0D	3844	0.8559 -	SFAM01.1	PPB	4/23/2024 11:29	SMPL
P2044-03	GCGT0S	15780	3.603 -	SFAM01.1	PPB	4/23/2024 11:32	SMPL
P2045-22DLX10	GCGY9	53472	12.2778 -	SFAM01.1	PPB	4/23/2024 11:38	SMPL
CCV054	CCV054	20787	4.7553 -	SFAM01.1	PPB	4/23/2024 11:45	SMPL
CCB054	CCB054	7	-0.0272 -	SFAM01.1	PPB	4/23/2024 11:48	SMPL
P2045-22DL2X25	GCGY9	21841	4.9979 -	SFAM01.1	PPB	4/23/2024 11:50	SMPL
CCV055	CCV055	20456	4.6791 -	SFAM01.1	PPB	4/23/2024 12:00	SMPL
CCB055	CCB055	85	-0.0093 -	SFAM01.1	PPB	4/23/2024 12:02	SMPL