

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130439

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/23/2024	09:38
2	CAL2	1	Water	NA	NA	20.3	7.00	04/23/2024	09:39
3	CAL3	1	Water	NA	NA	20.3	10.02	04/23/2024	09:40
4	ICV	1	Water	NA	NA	20.3	7.01	04/23/2024	09:45
5	CCV1	1	Water	NA	NA	20.3	2.01	04/23/2024	09:50
6	P2224-01	1	Water	NA	NA	20.7	7.06	04/23/2024	09:55
7	P2224-01DUP	1	Water	NA	NA	20.7	7.07	04/23/2024	09:56
8	CCV2	1	Water	NA	NA	20.3	12.02	04/23/2024	10:00

WORKLIST(Hardcopy Internal Chain)

130439

WorkList Name : ph p2224

WorkList ID : 179646

Department : Wet-Chemistry

Date : 04-23-2024 09:21:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2224-01	CLEANOUT-TOTE	Water	pH	Cool 4 deg C	PSEG03	I41	04/22/2024	9040C

Date/Time 04/23/24 09:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]