

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130441

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	04/23/2024	12:15
2	P2185-01		1	12	100.00	>212.0	>212.0	04/23/2024	12:45
3	P2185-01DUP		1	13	100.00	>212.0	>212.0	04/23/2024	13:15
4	P2186-03		1	12	100.00	>212.0	>212.0	04/23/2024	13:45
5	P2224-01		1	14	100.00	>212.0	>212.0	04/23/2024	14:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-130441

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 04/23/24 flashpoint

WorkList ID : 179644

Department : Wet-Chemistry

Date : 04-23-2024 08:15:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2185-01	TOTE-1	Water	Flash Point	Cool 4 deg C	PSEG03	I51	04/18/2024	1010B
P2186-03	NWB-2030	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/18/2024	1010B
P2224-01	CLEANOUT-TOTE	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/22/2024	1010B

Date/Time 04/23/2024 11:45

Raw Sample Received by: N.P.W.C

Raw Sample Relinquished by: JHCOX

Date/Time 04/23/2024 15:00

Raw Sample Received by: JHCOX

Raw Sample Relinquished by: N.P.W.C